

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111023\
 Data File : PD079524.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2023 20:34
 Operator : AR\AJ
 Sample : INDA4001
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA4007

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:21:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111023CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 08 01:20:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.569	2.804	54239135	85160939	38.145	38.781
27)	SA Decachlor...	9.118	7.950	130.5E6	183.8E6	76.893	76.751
Target Compounds							
2)	A alpha-BHC	4.026	3.307	94952067	141.2E6	41.240	41.082
3)	MA gamma-BHC...	4.360	3.636	90296237	131.0E6	40.668	40.763
4)	MA Heptachlor	4.953	3.979	93898728	128.1E6	39.237	40.127
9)	A Endosulfan I	6.109	5.132	76629537	107.8E6	38.183	39.546
13)	MA Dieldrin	6.385	5.397	169.2E6	249.1E6	80.811	82.467
14)	MA Endrin	6.615	5.672	139.4E6	216.8E6	80.031	81.754
16)	A 4,4'-DDD	6.746	5.819	125.7E6	189.2E6	80.974	82.440
17)	MA 4,4'-DDT	7.061	6.071	130.5E6	192.5E6	81.825	83.968
20)	A Methoxychlor	7.536	6.645	343.6E6	487.8E6	388.334	382.340

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111023\
Data File : PD079524.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2023 20:34
Operator : AR\AJ
Sample : INDA401
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:21:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111023CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Wed Nov 08 01:20:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

