

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111223\
 Data File : PD079550.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2023 13:47
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/09/2023
 Supervised By :Ankita Jodhani 11/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 05:11:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX102023.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 20 02:45:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.569	2.804	72865795	118.7E6	48.661	53.847
7) SA Decachlor...	9.117	7.948	91599123	126.2E6	48.819	55.211m
Target Compounds						
2) Toxaphene-1	6.277	5.036	9611822	9815268	490.923	549.824
3) Toxaphene-2	6.680	5.720	7641626	10234326	475.531	569.661
4) Toxaphene-3	7.100	6.759	30631564	42826952	478.901	577.432
5) Toxaphene-4	7.191	7.075	22521292	17429231	479.690	546.221
6) Toxaphene-5	7.612	7.207	16169382	12581074	471.764	561.791

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111223\
Data File : PD079550.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2023 13:47
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 05:11:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX102023.M
Quant Title : GC Extractables
QLast Update : Fri Oct 20 02:45:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm

