

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111319\
 Data File : PD055960.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2019 23:42
 Operator : SG\AJ
 Sample : K5774-09
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AR5

Manual Integrations
 APPROVED

Ankita
 11/15/2019 8:49:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 00:51:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 02 07:20:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.289	3.963	8561370	11249025	11.373	10.734
27)	SA Decachlor...	7.962	8.995	13210824	16430739	16.207	14.469m
Target Compounds							
3)	MA gamma-BHC...	3.997	4.611	4153205	4182506	4.060	2.969m#
12)	B 4,4'-DDE	5.413	6.282	1487511	1645953	1.711m	1.221m#
14)	MA Endrin	5.776	6.650	441025	719091	0.550m	0.625m
15)	B Endosulfa...	6.053	6.850	1968892	1279332	2.337m	0.988m#
16)	A 4,4'-DDD	5.913	6.770	2120052	4359246	3.002m	3.904m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111319\
Data File : PD055960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2019 23:42
Operator : SG\AJ
Sample : K5774-09
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0AR5

Manual Integrations
APPROVED

Ankita
11/15/2019 8:49:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 00:51:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 02 07:20:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

