

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111323\
 Data File : PD079595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2023 09:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/10/2023
 Supervised By :Ankita Jodhani 11/10/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 20:01:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 20 04:44:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.576	2.800	74785724	119.7E6	46.319	51.722
28)	SA Decachlor...	9.127	7.950	95222073	130.6E6	47.468	53.575
Target Compounds							
2)	A alpha-BHC	4.033	3.304	127.4E6	188.0E6	47.852	54.231
3)	MA gamma-BHC...	4.367	3.633	123.2E6	184.0E6	47.410	54.184
4)	MA Heptachlor	4.960	3.976	121.5E6	171.2E6	45.415	56.057
5)	MB Aldrin	5.304	4.257	121.7E6	178.7E6	47.217	54.641
6)	B beta-BHC	4.561	3.928	49778202	74945648	46.453	53.187
7)	B delta-BHC	4.810	4.160	126.5E6	187.4E6	47.254	55.962
8)	B Heptachlo...	5.731	4.759	111.2E6	160.8E6	47.429	54.248
9)	A Endosulfan I	6.117	5.129	102.8E6	143.2E6	47.405	54.068m
10)	B gamma-Chl...	5.986	5.008	109.4E6	156.8E6	47.695	54.650m
11)	B alpha-Chl...	6.067	5.073	108.0E6	154.1E6	47.380	53.695m
12)	B 4,4'-DDE	6.237	5.264	100.9E6	148.4E6	47.771	55.726
13)	MA Dieldrin	6.392	5.395	110.8E6	164.4E6	47.559	55.341
14)	MA Endrin	6.623	5.671	89336041	142.0E6	45.870	55.504
15)	B Endosulfa...	6.840	5.963	95423381	138.4E6	46.785	55.455
16)	A 4,4'-DDD	6.754	5.818	84096765	127.7E6	47.838	55.919
17)	MA 4,4'-DDT	7.069	6.070	86472893	126.4E6	45.664	55.202
18)	B Endrin al...	6.970	6.143	72276325	105.1E6	46.802	56.522
19)	B Endosulfa...	7.205	6.365	92586826	131.1E6	46.570	56.145
20)	A Methoxychlor	7.543	6.644	48490375	71714697	44.252	52.823
21)	B Endrin ke...	7.691	6.871	105.2E6	157.5E6	48.261	56.346
22)	Mirex	8.168	7.057	79058070	116.8E6	46.706	54.693

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111323\
 Data File : PD079595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2023 09:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 11/10/2023

Supervised By :Ankita Jodhani 11/10/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 20:01:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 20 04:44:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

