

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111419\
 Data File : PD055979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2019 10:52
 Operator : SG\AJ
 Sample : K5629-03MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C5AG1MSD

Manual Integrations
 APPROVED

Ankita
 11/15/2019 11:23:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 02:55:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 02 07:20:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.289	3.962	17322872	22255737	23.012	21.237
27)	SA Decachlor...	7.961	8.995	31197999	38448095	38.274	33.858
Target Compounds							
3)	MA gamma-BHC...	3.991	4.631	59268880	76775722	57.945	54.495
4)	MA Heptachlor	4.276	5.137	52966253	75757266	55.942	52.090
5)	MB Aldrin	4.515	5.440	52787062	72038427	56.977	52.502
8)	B Heptachlo...	4.952	5.819	3105710	8053706	3.478	5.968 #
9)	A Endosulfan I	5.269	6.156	2827812	3506255	3.374m	2.693m
10)	B trans-Chl...	5.173	6.055	2010959	3704414	2.218m	2.597m
11)	B cis-Chlor...	5.229	6.128	3423878	5708971	3.867m	4.068m
13)	MA Dieldrin	5.523	6.435	130.9E6	182.7E6	139.195	129.261
14)	MA Endrin	5.775	6.652	95915537	130.2E6	119.561	113.075
15)	B Endosulfa...	6.051	6.860	540909	1501442	0.642m	1.159m#
17)	MA 4,4'-DDT	6.145	7.071	79509906	116.0E6	122.696	106.713
21)	B Endrin ke...	6.911	7.673	4159430	6256683	4.696	4.744

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111419\
Data File : PD055979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2019 10:52
Operator : SG\AJ
Sample : K5629-03MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 02:55:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 02 07:20:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
11/15/2019 11:23:32 AM

Instrument :
ECD_D
Client Sampled :
CSAG1MSD