

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111422\
 Data File : PD072962.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2022 12:14
 Operator : AR\AJ
 Sample : PB148822BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS822

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 20:30:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110822CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 08 16:13:46 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.701	3.461	40515727	906.6E6	17.360	20.797
2)	SA Decachlor...	7.829	8.913	80556741	462.6E6	35.465	43.946
Target Compounds							
2)	A alpha-BHC	3.203	3.925	13833515	317.2E6	3.920	5.605 #
3)	MA gamma-BHC...	3.533	4.258	13391917	260.6E6	4.119	5.822 #
4)	MA Heptachlor	3.865	4.825	13650408	180.7E6	4.155	5.869 #
5)	MB Aldrin	4.144	5.163	13988725	154.3E6	4.323	5.881 #
6)	B beta-BHC	3.835	4.473	6630949	96965844	4.348	5.483 #
7)	B delta-BHC	4.064	4.717	13299182	198.1E6	3.824	5.527 #
8)	B Heptachlo...	4.649	5.596	12981464	128.5E6	4.333	5.806 #
9)	A Endosulfan I	5.018	5.980	12080643	109.0E6	4.365	5.648 #
10)	B trans-Chl...	4.897	5.849	12981810	118.2E6	4.391	5.729 #
11)	B cis-Chlor...	4.961	5.925	13032082	116.1E6	4.314	5.569 #
12)	B 4,4'-DDE	5.149	6.098	22652366	215.3E6	7.852	11.252 #
13)	MA Dieldrin	5.284	6.257	24444654	224.7E6	8.154	11.331 #
14)	MA Endrin	5.560	6.488	22169407	185.2E6	8.292	11.213 #
15)	B Endosulfa...	5.858	6.718	21956213	178.3E6	8.456	11.194 #
16)	A 4,4'-DDD	5.707	6.623	19093291	168.5E6	8.073	11.486 #
17)	MA 4,4'-DDT	5.956	6.931	17931984	159.7E6	7.759	11.224 #
18)	B Endrin al...	6.041	6.850	18571069	144.5E6	8.900	11.582 #
19)	B Endosulfa...	6.263	7.082	20797516	151.8E6	8.546	11.060 #
20)	A Methoxychlor	6.540	7.410	56317185	374.7E6	43.848	59.142 #
21)	B Endrin ke...	6.774	7.572	25211133	177.8E6	8.920	11.830 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111422\
Data File : PD072962.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2022 12:14
Operator : AR\AJ
Sample : PB148822BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 20:30:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110822CLP.M
Quant Title : GC Extractables
QLast Update : Tue Nov 08 16:13:46 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

