

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091216.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 02:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:50:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	161.9E6	1712.2E6	46.944	46.744
28)	SA Decachlor...	9.063	8.058	234.5E6	1929.3E6	47.066	48.750
Target Compounds							
2)	A alpha-BHC	3.993	3.385	389.5E6	2971.8E6	48.711	47.020
3)	MA gamma-BHC...	4.323	3.722	361.7E6	2705.5E6	48.407	46.790
4)	MA Heptachlor	4.921	4.073	342.3E6	2570.1E6	47.802	46.553
5)	MB Aldrin	5.262	4.358	346.2E6	2655.5E6	47.703	47.034
6)	B beta-BHC	4.511	4.017	134.5E6	1112.9E6	46.307	46.318
7)	B delta-BHC	4.759	4.254	363.7E6	2726.4E6	48.939	46.623
8)	B Heptachlo...	5.682	4.862	315.0E6	2379.8E6	48.485	46.700
9)	A Endosulfan I	6.066	5.236	289.6E6	2099.4E6	47.150	43.873
10)	B gamma-Chl...	5.938	5.114	305.6E6	2572.8E6	47.002	47.097
11)	B alpha-Chl...	6.018	5.179	325.4E6	2447.4E6	47.270	46.767
12)	B 4,4'-DDE	6.189	5.363	275.2E6	2518.6E6	47.460	47.127
13)	MA Dieldrin	6.339	5.501	308.1E6	2539.6E6	47.654	46.761
14)	MA Endrin	6.567	5.778	238.3E6	2112.7E6	48.004	46.749
15)	B Endosulfa...	6.780	6.069	265.7E6	2177.2E6	47.801	46.642
16)	A 4,4'-DDD	6.699	5.918	239.5E6	2274.3E6	48.761	46.539
17)	MA 4,4'-DDT	7.014	6.171	200.4E6	1885.4E6	48.454	48.180
18)	B Endrin al...	6.908	6.248	195.9E6	1637.4E6	47.032	46.401
19)	B Endosulfa...	7.143	6.471	241.2E6	2103.5E6	46.874	46.918
20)	A Methoxychlor	7.487	6.742	103.5E6	965.5E6	47.582	48.149
21)	B Endrin ke...	7.624	6.980	271.0E6	2430.9E6	48.096	47.706
22)	Mirex	8.106	7.173	166.7E6	1571.9E6	46.878	47.431

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091216.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 02:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:50:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

