

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111618\
 Data File : PD050371.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Nov 2018 4:57
 Operator : AJ\SJ
 Sample : PB114852TB
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 06:18:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111618.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 16 02:59:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.228	4.044	48275448	887.9E6	33.174	32.238
28)	SA Decachlor...	7.838	9.154	54598976	641.9E6	36.883	31.711
Target Compounds							
2)	A alpha-BHC	3.672	4.504f	812169	180.2E6	0.346	4.246 #
3)	MA gamma-BHC...	3.923	0.000	1473819	0	0.632	N.D. #
6)	B beta-BHC	4.105f	4.942f	332803	37192866	0.364	2.242 #
13)	MA Dieldrin	0.000	6.554	0	2865324	N.D.	0.095 #
15)	B Endosulfa...	0.000	6.979	0	33908716	N.D.	1.389 #
19)	B Endosulfa...	6.242f	0.000	1861909	0	1.004	N.D. #
20)	A Methoxychlor	0.000	7.629	0	7344743	N.D.	0.661 #
21)	B Endrin ke...	0.000	7.786	0	5180381	N.D.	0.224 #
24)	Chlordane-2	0.000	5.473	0	7311808	N.D.	5.861 #
27)	Chlordane-5	0.000	7.048	0	4969225	N.D.	5.438 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111618\
Data File : PD050371.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Nov 2018 4:57
Operator : AJ\SJ
Sample : PB114852TB
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 06:18:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111618.M
Quant Title : GC Extractables
QLast Update : Fri Nov 16 02:59:21 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.228 min
Delta R.T.: 0.000 min
Response: 48275448
Conc: 33.17 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: 0.000 min
Response: 887946296
Conc: 32.24 ng/ml




#2 alpha-BHC

R.T.: 3.672 min
Delta R.T.: 0.023 min
Response: 812169
Conc: 0.35 ng/ml



#2 alpha-BHC


R.T.: 4.504 min
Delta R.T.: 0.068 min
Response: 180219476
Conc: 4.25 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.923 min
Delta R.T.: 0.005 min
Response: 1473819
Conc: 0.63 ng/ml

Instrument :
ECD_D
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 4.723 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.105 min
Delta R.T.: -0.058 min
Response: 332803
Conc: 0.36 ng/ml



#6 beta-BHC


R.T.: 4.942 min
Delta R.T.: 0.050 min
Response: 37192866
Conc: 2.24 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T. : 5.416 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#13 Dieldrin

R.T.: 6.554 min
Delta R.T.: 0.005 min
Response: 2865324
Conc: 0.10 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 5.933 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.979 min
Delta R.T.: 0.003 min
Response: 33908716
Conc: 1.39 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.242 min
Delta R.T.: -0.065 min
Response: 1861909
Conc: 1.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 7.324 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.582 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.629 min
Delta R.T.: -0.016 min
Response: 7344743
Conc: 0.66 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T.: 6.783 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#21 Endrin ketone

R.T.: 7.786 min
Delta R.T.: -0.011 min
Response: 5180381
Conc: 0.22 ng/ml




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T.: 4.533 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 5.473 min
Delta R.T.: -0.045 min
Response: 7311808
Conc: 5.86 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 5.937 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#27 Chlordane-5

R.T.: 7.048 min
Delta R.T.: -0.002 min
Response: 4969225
Conc: 5.44 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.838 min
Delta R.T.: 0.002 min
Response: 54598976
Conc: 36.88 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.154 min
Delta R.T.: 0.002 min
Response: 641893744
Conc: 31.71 ng/ml