

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111622\
 Data File : PD073012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Nov 2022 16:10
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 17 03:14:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111622.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 17 03:13:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.699	3.460	107.0E6	2741.4E6	50.000	50.000
28)	SA Decachlor...	7.829	8.913	101.1E6	635.7E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.202	3.924	180.0E6	3619.6E6	50.000	50.000
3)	MA gamma-BHC...	3.532	4.258	144.0E6	2754.3E6	50.000	50.000
4)	MA Heptachlor	3.865	4.825	150.9E6	1831.6E6	50.000	50.000
5)	MB Aldrin	4.144	5.163	153.0E6	1573.8E6	50.000	50.000
6)	B beta-BHC	3.834	4.473	63504087	1020.2E6	50.000	50.000
7)	B delta-BHC	4.064	4.717	160.9E6	1975.1E6	50.000	50.000
8)	B Heptachlo...	4.649	5.595	135.7E6	1305.7E6	50.000	50.000
9)	A Endosulfan I	5.018	5.980	127.1E6	1203.9E6	50.000	50.000
10)	B gamma-Chl...	4.897	5.849	139.0E6	1327.8E6	50.000	50.000
11)	B alpha-Chl...	4.961	5.925	136.2E6	1268.7E6	50.000	50.000
12)	B 4,4'-DDE	5.149	6.098	130.1E6	1228.4E6	50.000	50.000
13)	MA Dieldrin	5.283	6.257	134.7E6	1245.0E6	50.000	50.000
14)	MA Endrin	5.560	6.488	122.6E6	1047.5E6	50.000	50.000
15)	B Endosulfa...	5.859	6.718	108.8E6	964.7E6	50.000	50.000
16)	A 4,4'-DDD	5.707	6.622	113.4E6	967.4E6	50.000	50.000
17)	MA 4,4'-DDT	5.956	6.930	100.5E6	947.0E6	50.000	50.000
18)	B Endrin al...	6.040	6.849	85880752	735.3E6	50.000	50.000
19)	B Endosulfa...	6.263	7.082	110.2E6	867.2E6	50.000	50.000
20)	A Methoxychlor	6.539	7.409	58011720	440.9E6	50.000	50.000
21)	B Endrin ke...	6.774	7.572	127.9E6	922.6E6	50.000	50.000
22)	Mirex	6.938	8.019	90557760	591.7E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111622\
Data File : PD073012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Nov 2022 16:10
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 17 03:14:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111622.M
Quant Title : GC Extractables
QLast Update : Thu Nov 17 03:13:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

