

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111718\
 Data File : PD050407.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Nov 2018 16:17
 Operator : AJ\SJ
 Sample : J5762-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BU-03-111518

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 17 01:19:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111618.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 16 02:59:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.230	4.043	13927406	287.3E6	9.571	10.430
28) SA Decachlor...	7.840	9.156	14838137	166.1E6	10.023	8.204
Target Compounds						
12) B 4,4'-DDE	5.304	6.396	4770003	86246116	2.327	2.933 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111718\
Data File : PD050407.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Nov 2018 16:17
Operator : AJ\SJ
Sample : J5762-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BU-03-111518

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 17 01:19:11 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111618.M
Quant Title : GC Extractables
QLast Update : Fri Nov 16 02:59:21 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.230 min
Delta R.T.: 0.002 min
Response: 13927406
Conc: 9.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
BU-03-111518




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 287284673
Conc: 10.43 ng/ml




#12 4,4'-DDE

R.T.: 5.304 min
Delta R.T.: 0.002 min
Response: 4770003
Conc: 2.33 ng/ml



#12 4,4'-DDE


R.T.: 6.396 min
Delta R.T.: 0.000 min
Response: 86246116
Conc: 2.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.840 min
Delta R.T.: 0.003 min
Response: 14838137
Conc: 10.02 ng/ml

Instrument :
ECD_D
ClientSampleId :
BU-03-111518



#28 Decachlorobiphenyl

R.T.: 9.156 min
Delta R.T.: 0.004 min
Response: 166056543
Conc: 8.20 ng/ml