

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111720\
 Data File : PD060093.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2020 15:29
 Operator : DD\AJ
 Sample : INDA316
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA316

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 06:41:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 04 04:05:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.506	4.023	16803371	61213137	19.608	20.535
27)	SA Decachlor...	8.306	9.110	30955389	91287599	37.256	37.108
Target Compounds							
2)	A alpha-BHC	3.948	4.414	25644609	89198234	19.955	20.025
3)	MA gamma-BHC...	4.234	4.699	23817799	82156130	19.485	20.300
4)	MA Heptachlor	4.541	5.212	22482436	79275032	19.155	19.995
9)	A Endosulfan I	5.583	6.263	19807915	63731506	19.474	19.885
13)	MA Dieldrin	5.829	6.521	41393125	139.7E6	37.991	40.001
14)	MA Endrin	6.088	6.739	33437946	109.7E6	37.065	39.178
16)	A 4,4'-DDD	6.213	6.859	32414277	107.9E6	36.653	40.203
17)	MA 4,4'-DDT	6.456	7.158	33708483	104.9E6	37.505	38.554
20)	A Methoxychlor	7.005	7.615	72302078	260.6E6	169.895	185.378

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111720\
Data File : PD060093.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2020 15:29
Operator : DD\AJ
Sample : INDA316
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 06:41:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Nov 04 04:05:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
INDA316