

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111723\
 Data File : PD079666.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2023 22:26
 Operator : AR\AJ
 Sample : 05328-15MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 A49W2MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 00:29:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 04 07:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.569	2.804	21730781	33990183	14.880	14.860
2)	SA Decachlor...	9.118	7.950	31828110	43900639	17.936	18.388
Target Compounds							
2)	A alpha-BHC	4.026	3.307	101.3E6	153.1E6	40.881	42.236
3)	MA gamma-BHC...	4.359	3.637	93055109	139.2E6	39.386	40.521
4)	MA Heptachlor	4.952	3.979	91277527	129.3E6	36.443	38.755
5)	MB Aldrin	5.296	4.260	86854377	124.9E6	37.038	38.394
6)	B beta-BHC	4.552	3.931	39288057	57628066	36.183	35.189
7)	B delta-BHC	4.802	4.163	95754589	142.0E6	38.426	39.734
8)	B Heptachlo...	5.722	4.761	85244986	124.2E6	38.452	40.143
9)	A Endosulfan I	6.108	5.132	80698001	113.1E6	39.037	40.781
10)	B trans-Chl...	5.977	5.012	81861413	120.6E6	38.474	40.630
11)	B cis-Chlor...	6.058	5.076	83601380	120.3E6	38.385	39.620
12)	B 4,4'-DDE	6.228	5.265	158.9E6	239.7E6	80.489	85.595
13)	MA Dieldrin	6.384	5.396	175.0E6	261.8E6	79.536	82.225
14)	MA Endrin	6.614	5.671	145.3E6	231.5E6	79.642	84.509
15)	B Endosulfa...	6.832	5.964	149.7E6	222.8E6	78.703	85.785
16)	A 4,4'-DDD	6.745	5.819	131.1E6	202.5E6	80.924	85.999
17)	MA 4,4'-DDT	7.060	6.070	136.7E6	201.1E6	80.227	83.707
18)	B Endrin al...	6.962	6.143	109.8E6	162.1E6	73.690	80.124
19)	B Endosulfa...	7.197	6.365	143.2E6	207.4E6	76.902	83.466
20)	A Methoxychlor	7.534	6.645	345.7E6	493.2E6	375.038	383.669
21)	B Endrin ke...	7.684	6.872	155.5E6	241.2E6	79.613	84.843

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111723\
Data File : PD079666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2023 22:26
Operator : AR\AJ
Sample : 05328-15MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 00:29:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 04 07:14:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

