

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111924\
 Data File : PD086789.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Nov 2024 08:58
 Operator : AR\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TOXAPH3159

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 00:06:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111424CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 14 03:38:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	


System Monitoring Compounds								
1)	SA Tetrachlo...	3.589f	2.917f	72376	542556	<MDL	<MDL	#
27)	SA Decachlor...	0.000	8.106	0	112472	N.D.	<MDL	
Target Compounds								
3)	MA gamma-BHC...	0.000	3.703f	0	145556	N.D.	<MDL	
4)	MA Heptachlor	4.935	0.000	14719	0	<MDL	N.D.	#
5)	MB Aldrin	5.289	0.000	46092	0	<MDL	N.D.	#
6)	B beta-BHC	4.509	4.001f	37933	359198	<MDL	<MDL	#
7)	B delta-BHC	4.795	0.000	152914	0	<MDL	N.D.	#
8)	B Heptachlo...	5.699	0.000	1730422	0	0.682	N.D.	#
9)	A Endosulfan I	6.072	0.000	2944	0	<MDL	N.D.	#
10)	B trans-Chl...	5.955	5.085f	145213	2771632	<MDL	0.191	#
11)	B cis-Chlor...	6.014	0.000	154251	0	<MDL	N.D.	#
12)	B 4,4'-DDE	6.216	5.339f	64556	297436	<MDL	<MDL	#
14)	MA Endrin	0.000	5.806	0	3594699	N.D.	0.280	
15)	B Endosulfa...	6.805	6.078	2647439	546751	1.109	<MDL	#
16)	A 4,4'-DDD	0.000	5.959	0	425840	N.D.	<MDL	
17)	MA 4,4'-DDT	7.080f	6.197	5332	111777	<MDL	<MDL	#
18)	B Endrin al...	0.000	6.307f	0	605718	N.D.	<MDL	
19)	B Endosulfa...	0.000	6.504	0	640796	N.D.	<MDL	
20)	A Methoxychlor	0.000	6.777	0	593561	N.D.	0.108	
22)	Toxaphene-1	6.216f	0.000	64556	0	4.102	N.D.	#
23)	Toxaphene-2	6.459	5.806f	442561	3594699	19.840	45.466	#
24)	Toxaphene-3	7.080	6.777	5332	593561	<MDL	2.294	#
25)	Toxaphene-4	0.000	7.217	0	686457	N.D.	3.782	
26)	Toxaphene-5	0.000	7.338	0	1485692	N.D.	10.861	


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111924\
Data File : PD086789.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Nov 2024 08:58
Operator : AR\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 00:06:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111424CLP.M
Quant Title : GC Extractables
QLast Update : Thu Nov 14 03:38:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.589 min
Delta R.T.: 0.032 min
Response: 72376
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#1 Tetrachloro-m-xylene

R.T.: 2.917 min
Delta R.T.: 0.032 min
Response: 542556
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.699 min
Delta R.T.: -0.001 min
Response: 1730422
Conc: 0.68 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.884 min
Response: 0
Conc: N.D.



#10 trans-Chlordane


R.T.: 5.955 min
Delta R.T.: 0.000 min
Response: 145213
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#10 trans-Chlordane

R.T.: 5.085 min
Delta R.T.: -0.053 min
Response: 2771632
Conc: 0.19 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.584 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.806 min
Delta R.T.: 0.005 min
Response: 3594699
Conc: 0.28 ng/ml



#15 Endosulfan II


R.T.: 6.805 min
Delta R.T.: 0.010 min
Response: 2647439
Conc: 1.11 ng/ml

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#15 Endosulfan II

R.T.: 6.078 min
Delta R.T.: -0.014 min
Response: 546751
Conc: N.D.




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 7.502 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 6.777 min
Delta R.T.: 0.007 min
Response: 593561
Conc: 0.11 ng/ml



#22 Toxaphene-1


R.T.: 6.216 min
Delta R.T.: -0.036 min
Response: 64556
Conc: 4.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T. : 5.159 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 6.459 min
Delta R.T.: 0.008 min
Response: 442561
Conc: 19.84 ng/ml



#23 Toxaphene-2


R.T.: 5.806 min
Delta R.T.: -0.043 min
Response: 3594699
Conc: 45.47 ng/ml



#24 Toxaphene-3


R.T.: 7.080 min
Delta R.T.: 0.014 min
Response: 5332
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#24 Toxaphene-3

R.T.: 6.777 min
Delta R.T.: 0.008 min
Response: 593561
Conc: 2.29 ng/ml




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T. : 7.158 min
Response: 0
Conc: N.D.



#25 Toxaphene-4


R.T.: 7.217 min
Delta R.T.: 0.006 min
Response: 686457
Conc: 3.78 ng/ml



#26 Toxaphene-5


R.T.: 0.000 min
Exp R.T.: 7.939 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
TOXAPH3159




#26 Toxaphene-5

R.T.: 7.338 min
Delta R.T.: -0.005 min
Response: 1485692
Conc: 10.86 ng/ml



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 9.087 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 8.106 min
Delta R.T.: 0.016 min
Response: 112472
Conc: N.D.