

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112020\
Data File : PD060156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2020 12:13
Operator : DD\AJ
Sample : L4711-14
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C0B66

Manual Integrations
APPROVED

mohammad
11/23/2020 12:03:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 07:04:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Nov 04 04:05:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(13) Dieldrin (MA)
5.825min 8.085 ng/ml
response 8809608

(13) Dieldrin #2 (MA)
6.519min 8.017 ng/ml
response 27995849

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112020\
Data File : PD060156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2020 12:13
Operator : DD\AJ
Sample : L4711-14
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C0B66

Manual Integrations
APPROVED

mohammad
11/23/2020 12:03:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 07:04:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Nov 04 04:05:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(13) Dieldrin (MA)
5.823min 9.450 ng/ml m
response 10295966

(13) Dieldrin #2 (MA)
6.519min 8.017 ng/ml
response 27995849

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112020\
 Data File : PD060156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2020 12:13
 Operator : DD\AJ
 Sample : L4711-14
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0B66

Manual Integrations
 APPROVED

mohammad
 11/23/2020 12:03:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 07:04:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 04 04:05:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.504	4.023	7541307	28392129	8.800	9.525
27) SA Decachlor...	8.301	9.107	8900954	24781305	10.713	10.074
Target Compounds						
10) B trans-Chl...	5.453	6.135	10665924	15797556	10.042	4.598 #
11) B cis-Chlor...	5.524	6.213	5381948	29463171	5.088	8.581 #
12) B 4,4'-DDE	5.689	6.366	7571706	22305114	7.408	6.898
13) MA Dieldrin	5.823	6.519	10295966	27995849	9.450m	8.017
15) B Endosulfa...	6.368	6.953	1852615	5808635	2.054	2.055
17) MA 4,4'-DDT	6.452	7.156	5762269	14529540	6.411	5.339

AJ
 11/25/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.