

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112023\
 Data File : PD079721.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 15:17
 Operator : AR\AJ
 Sample : PIBLK086
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 20:23:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 04 07:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.568	2.803	25795198	41405553	17.663	18.102
27)	SA Decachlor...	9.113	7.947	63633192	83014016	35.859	34.771
Target Compounds							
2)	A alpha-BHC	0.000	3.325	0	130216	N.D.	<MDL
3)	MA gamma-BHC...	0.000	3.669f	0	400019	N.D.	0.116
4)	MA Heptachlor	0.000	3.942f	0	151553	N.D.	<MDL
5)	MB Aldrin	0.000	4.279	0	307984	N.D.	<MDL
6)	B beta-BHC	0.000	3.942	0	151553	N.D.	<MDL
9)	A Endosulfan I	0.000	5.096	0	130385	N.D.	<MDL
11)	B cis-Chlor...	0.000	5.096	0	130385	N.D.	<MDL
13)	MA Dieldrin	0.000	5.395	0	575466	N.D.	0.181
14)	MA Endrin	0.000	5.686	0	504305	N.D.	0.184
15)	B Endosulfa...	6.802	5.921f	1960160	1343935	1.030	0.517 #
16)	A 4,4'-DDD	6.802f	5.823	1960160	27735	1.210	<MDL #
18)	B Endrin al...	0.000	6.156	0	371979	N.D.	0.184
19)	B Endosulfa...	0.000	6.414f	0	101799	N.D.	<MDL
21)	B Endrin ke...	0.000	6.888	0	98819	N.D.	<MDL
22)	Toxaphene-1	0.000	5.096f	0	130385	N.D.	7.798
23)	Toxaphene-2	0.000	5.395	0	575466	N.D.	34.245
25)	Toxaphene-4	0.000	6.718f	0	318450	N.D.	4.046


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112023\
Data File : PD079721.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 15:17
Operator : AR\AJ
Sample : PIBLK086
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 20:23:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 04 07:14:58 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.568 min
Delta R.T.: 0.001 min
Response: 25795198
Conc: 17.66 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.803 min
Delta R.T.: 0.000 min
Response: 41405553
Conc: 18.10 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 4.357 min
Response: 0
Conc: N.D.




#3 gamma-BHC (Lindane)

R.T.: 3.669 min
Delta R.T.: 0.034 min
Response: 400019
Conc: 0.12 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 6.382 min
Response: 0
Conc: N.D.




#13 Dieldrin

R.T.: 5.395 min
Delta R.T.: 0.000 min
Response: 575466
Conc: 0.18 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.612 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 5.686 min
Delta R.T.: 0.015 min
Response: 504305
Conc: 0.18 ng/ml




#15 Endosulfan II

R.T.: 6.802 min
Delta R.T.: -0.027 min
Response: 1960160
Conc: 1.03 ng/ml




#15 Endosulfan II

R.T.: 5.921 min
Delta R.T.: -0.042 min
Response: 1343935
Conc: 0.52 ng/ml




#16 4,4'-DDD

R.T.: 6.802 min
Delta R.T.: 0.059 min
Response: 1960160
Conc: 1.21 ng/ml




#16 4,4'-DDD

R.T.: 5.823 min
Delta R.T.: 0.005 min
Response: 27735
Conc: N.D.




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.959 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 6.156 min
Delta R.T.: 0.013 min
Response: 371979
Conc: 0.18 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T. : 5.886 min
Response: 0
Conc: N.D.




#22 Toxaphene-1

R.T.: 5.096 min
Delta R.T.: 0.060 min
Response: 130385
Conc: 7.80 ng/ml




#23 Toxaphene-2

R.T.: 0.000 min
Exp R.T.: 6.477 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 5.395 min
Delta R.T.: 0.034 min
Response: 575466
Conc: 34.24 ng/ml




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T.: 7.186 min
Response: 0
Conc: N.D.




#25 Toxaphene-4

R.T.: 6.718 min
Delta R.T.: -0.041 min
Response: 318450
Conc: 4.05 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.113 min
Delta R.T.: 0.003 min
Response: 63633192
Conc: 35.86 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.947 min
Delta R.T.: -0.002 min
Response: 83014016
Conc: 34.77 ng/ml