

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112023\
 Data File : PD079733.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 19:18
 Operator : AR\AJ
 Sample : 05328-15MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 A49W2MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 20:27:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 04 07:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.568	2.804	21297899	33088914	14.584	14.466
2)	SA Decachlor...	9.112	7.946	23096549	33159242	13.015	13.889
Target Compounds							
2)	A alpha-BHC	4.025	3.306	98858644	153.3E6	39.900	42.281
3)	MA gamma-BHC...	4.358	3.636	89598608	139.0E6	37.923	40.467
4)	MA Heptachlor	4.951	3.978	86588113	128.8E6	34.571	38.609
5)	MB Aldrin	5.295	4.259	80364059	122.7E6	34.270	37.713
6)	B beta-BHC	4.552	3.930	37124639	56982814	34.191	34.795
7)	B delta-BHC	4.801	4.162	90875756	141.3E6	36.468	39.561
8)	B Heptachlo...	5.720	4.760	79050589	121.0E6	35.658	39.116
9)	A Endosulfan I	6.107	5.130	75322632	109.8E6	36.437	39.596
10)	B trans-Chl...	5.976	5.011	76931129	117.1E6	36.157	39.435
11)	B cis-Chlor...	6.056	5.075	78260945	116.6E6	35.933	38.427
12)	B 4,4'-DDE	6.227	5.263	149.2E6	231.7E6	75.549	82.747
13)	MA Dieldrin	6.382	5.394	162.8E6	252.6E6	73.999	79.343
14)	MA Endrin	6.613	5.670	136.2E6	221.4E6	74.670	80.824
15)	B Endosulfa...	6.830	5.962	137.4E6	207.5E6	72.210	79.898
16)	A 4,4'-DDD	6.743	5.817	118.9E6	190.1E6	73.399	80.737
17)	MA 4,4'-DDT	7.059	6.068	123.0E6	191.8E6	72.172	79.836
18)	B Endrin al...	6.960	6.142	98672674	152.4E6	66.201	75.303
19)	B Endosulfa...	7.195	6.363	126.0E6	194.6E6	67.686	78.313
20)	A Methoxychlor	7.532	6.643	315.2E6	474.2E6	341.944	368.947
21)	B Endrin ke...	7.681	6.869	131.6E6	218.5E6	67.354	76.843

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112023\
Data File : PD079733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 19:18
Operator : AR\AJ
Sample : 05328-15MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 20:27:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 04 07:14:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

