

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112025\
 Data File : PD091326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2025 20:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 22:51:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.872	169.1E6	1879.2E6	49.048	51.303
28)	SA Decachlor...	9.058	8.055	228.6E6	1869.8E6	45.879	47.246
Target Compounds							
2)	A alpha-BHC	3.990	3.383	410.5E6	3290.0E6	51.344	52.055
3)	MA gamma-BHC...	4.321	3.719	380.5E6	3000.1E6	50.919	51.885
4)	MA Heptachlor	4.918	4.070	355.6E6	2839.0E6	49.648	51.423
5)	MB Aldrin	5.259	4.355	358.5E6	2944.8E6	49.404	52.160
6)	B beta-BHC	4.508	4.015	140.9E6	1209.0E6	48.496	50.316
7)	B delta-BHC	4.757	4.251	366.1E6	2998.4E6	49.255	51.276
8)	B Heptachlo...	5.679	4.859	315.2E6	2566.6E6	48.521	50.366
9)	A Endosulfan I	6.063	5.233	295.6E6	2199.3E6	48.116	45.962
10)	B gamma-Chl...	5.936	5.112	316.1E6	2805.1E6	48.630	51.350
11)	B alpha-Chl...	6.016	5.176	317.6E6	2655.5E6	46.139	50.744
12)	B 4,4'-DDE	6.185	5.361	279.5E6	2771.8E6	48.191	51.865
13)	MA Dieldrin	6.336	5.499	314.3E6	2718.2E6	48.617	50.049
14)	MA Endrin	6.563	5.775	260.0E6	2436.0E6	52.385	53.903
15)	B Endosulfa...	6.776	6.067	264.9E6	2266.1E6	47.653	48.547
16)	A 4,4'-DDD	6.696	5.915	229.3E6	2271.6E6	46.683	46.484
17)	MA 4,4'-DDT	7.011	6.168	211.1E6	2085.1E6	51.042	53.284
18)	B Endrin al...	6.904	6.245	190.2E6	1670.8E6	45.664	47.346
19)	B Endosulfa...	7.139	6.468	241.0E6	2169.8E6	46.851	48.395
20)	A Methoxychlor	7.484	6.739	110.0E6	1029.2E6	50.608	51.329
21)	B Endrin ke...	7.621	6.977	267.6E6	2382.8E6	47.501	46.763
22)	Mirex	8.102	7.169	162.3E6	1565.1E6	45.619	47.228

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112025\
Data File : PD091326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2025 20:04
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 22:51:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

