

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112222\
 Data File : PD073187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 17:01
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORCCC500

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/23/2022
 Supervised By : Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 20:45:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111622.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 17 03:26:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.703	3.458	124.9E6	2960.8E6	58.858	53.443
28)	SA Decachlor...	7.830	8.909	105.9E6	694.4E6	51.419	54.327
Target Compounds							
23)	Chlordane-1	3.695	4.606	41473693	790.4E6	499.482	504.562
24)	Chlordane-2	4.271	5.137	50516846	562.8E6	515.339	458.784m
25)	Chlordane-3	4.901	5.845	151.5E6	2065.4E6	519.560	541.956
26)	Chlordane-4	4.963	5.922	153.4E6	2388.1E6	527.476	541.402
27)	Chlordane-5	5.860	6.780	43853787	423.4E6	453.934	571.537 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112222\
Data File : PD073187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 17:01
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/23/2022
Supervised By : Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 20:45:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111622.M
Quant Title : GC Extractables
QLast Update : Thu Nov 17 03:26:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

