

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112320\
 Data File : PD060273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Nov 2020 21:50
 Operator : DD\AJ
 Sample : L4751-01
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0B79

Manual Integrations
 APPROVED

Ankita
 11/25/2020 5:31:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 00:39:31 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 04 04:05:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.504	4.023	9421353	26966387	10.994	9.046
27)	SA Decachlor...	8.291	9.107	24369289	62307787	29.329	25.328m
Target Compounds							
3)	MA gamma-BHC...	4.235	4.678	9509255	33155756	7.779m	8.192m
10)	B trans-Chl...	5.465	6.122	3153260	203.9E6	2.969m	59.359 #
11)	B cis-Chlor...	5.523	6.212	3998056	47164856	3.779m	13.737m#
12)	B 4,4'-DDE	5.687	6.363	1513404	11824304	1.481m	3.657 #
16)	A 4,4'-DDD	6.206	6.854	5392949	41460044	6.098m	15.455 #
17)	MA 4,4'-DDT	6.448	7.160	4850276	3138319	5.397m	1.153 #
19)	B Endosulfa...	6.757	7.289	7101644	5114572	8.126	1.774m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112320\
Data File : PD060273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Nov 2020 21:50
Operator : DD\AJ
Sample : L4751-01
Misc :
ALS Vial : 52 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 00:39:31 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Nov 04 04:05:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
11/25/2020 5:31:30 PM

Instrument :
ECD_D
ClientSample :
C0879