

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112421\
 Data File : PD067081.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2021 10:46
 Operator : AR\AJ
 Sample : PB140981BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS981

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 03:00:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 16:50:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.437	3.966	34440386	182.9E6	17.816	21.119
27)	SA Decachlor...	8.161	9.024	36374984	203.2E6	31.045	38.962 #
Target Compounds							
2)	A alpha-BHC	3.873	4.359	11286320	53606839	3.856	4.903 #
3)	MA gamma-BHC...	4.155	4.644	11406819	51543528	4.120	5.100
4)	MA Heptachlor	4.448	5.144	12206868	49191590	4.291	5.019
5)	MB Aldrin	4.693	5.447	12106426	45653518	4.205	4.845
6)	B beta-BHC	4.403	4.821	5783189	26361303	4.604	5.350
7)	B delta-BHC	4.605	5.036	3940694	23535606	1.487	2.379 #
8)	B Heptachlo...	5.133	5.836	11528366	33716001	4.535	4.485
9)	A Endosulfan I	5.469	6.194	10534862	40702423	4.596	5.202
10)	B trans-Chl...	5.357	6.071	11535954	42093929	4.463	4.888
11)	B cis-Chlor...	5.415	6.142	11483249	43143320	4.507	5.028
12)	B 4,4'-DDE	5.580	6.302	20083497	77327736	8.793	10.155
13)	MA Dieldrin	5.711	6.451	21944354	81292582	8.794	10.190
14)	MA Endrin	5.969	6.670	18820316	68104944	8.895	10.036
15)	B Endosulfa...	6.241	6.885	18985398	74060755	9.143	10.328
16)	A 4,4'-DDD	6.093	6.797	16030186	65738659	8.856	10.183
17)	MA 4,4'-DDT	6.333	7.094	16166700	62364989	8.550	9.440
18)	B Endrin al...	6.410	7.010	16257488	61716284	11.159	11.600
19)	B Endosulfa...	6.622	7.233	14630336	57712901	8.021	8.374
20)	A Methoxychlor	6.880	7.557	43371675	156.3E6	46.277	48.351
21)	B Endrin ke...	7.114	7.706	19161451	74348145	9.299	9.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112421\
Data File : PD067081.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2021 10:46
Operator : AR\AJ
Sample : PB140981BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 03:00:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 27 16:50:52 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

