

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112619\
 Data File : PD056197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2019 15:01
 Operator : SG\AJ
 Sample : PB125111BL
 Misc : 608
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 00:29:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 02 07:20:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.294	3.960	15845934	23323511	21.050	22.256
27)	SA Decachlor...	7.972	8.999	19423312	24951715	23.829	21.973
Target Compounds							
14)	MA Endrin	0.000	6.667	0	152914	N.D.	0.133
15)	B Endosulfa...	0.000	6.856	0	728403	N.D.	0.562
16)	A 4,4'-DDD	0.000	6.712f	0	3070395	N.D.	2.750


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112619\
Data File : PD056197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2019 15:01
Operator : SG\AJ
Sample : PB125111BL
Misc : 608
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 00:29:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110319CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 02 07:20:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.294 min
Delta R.T.: 0.005 min
Response: 15845934
Conc: 21.05 ng/ml

Instrument :
ECD_D
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 3.960 min
Delta R.T.: -0.002 min
Response: 23323511
Conc: 22.26 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 5.776 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 6.667 min
Delta R.T.: 0.015 min
Response: 152914
Conc: 0.13 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.049 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
HEXANE




#15 Endosulfan II

R.T.: 6.856 min
Delta R.T.: -0.004 min
Response: 728403
Conc: 0.56 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 5.911 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.712 min
Delta R.T.: -0.062 min
Response: 3070395
Conc: 2.75 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.972 min
Delta R.T.: 0.009 min
Response: 19423312
Conc: 23.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
HEXANE



#27 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: 0.002 min
Response: 24951715
Conc: 21.97 ng/ml