

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112921\
Data File : PD067108.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2021 17:01
Operator : AR\AJ
Sample : PEM048
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
PEM048

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/30/2021
Supervised By :mohammad ahmed 11/30/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 05:31:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
Quant Title : GC Extractables
QLast Update : Tue Nov 30 05:28:01 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)

0.000min 0.000 ng/ml

response 0

(12) 4,4'-DDE #2 (B)

6.303min 0.382 ng/ml

response 2562540

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112921\
 Data File : PD067108.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2021 17:01
 Operator : AR\AJ
 Sample : PEM048
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 ECD_D
 LabSampleId :
 PEM048

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/30/2021
 Supervised By :mohammad ahmed 11/30/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 05:31:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 05:28:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)
 5.578min 0.276 ng/ml m
 response 690411

(12) 4,4'-DDE #2 (B)
 6.303min 0.382 ng/ml
 response 2562540

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112921\
 Data File : PD067108.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2021 17:01
 Operator : AR\AJ
 Sample : PEM048
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 LabSampleId :
 PEM048

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 11/30/2021
 Supervised By : mohammad ahmed 11/30/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 05:31:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 05:28:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.437	3.966	39190974	145.9E6	17.949	18.209
27) SA Decachlor...	8.162	9.022	22589986	98966641	19.897	19.517
Target Compounds						
2) A alpha-BHC	3.873	4.360	29691486	104.0E6	9.078	9.697
3) MA gamma-BHC...	4.155	4.645	28779245	93981702	9.043	9.369
6) B beta-BHC	4.403	4.822	13417622	48580498	10.079	10.530
12) B 4,4'-DDE	5.578	6.303	690411	2562540	0.276m	0.382 #
14) MA Endrin	5.969	6.672	106.6E6	270.0E6	46.621	46.530
16) A 4,4'-DDD	6.094	6.797	2919147	8687507	1.381	1.486
17) MA 4,4'-DDT	6.333	7.095	200.3E6	510.1E6	94.228	88.518
18) B Endrin al...	6.410	7.011	3087424	6108377	1.872	1.296 #
20) A Methoxychlor	6.881	7.557	225.6E6	624.4E6	227.063	224.357
21) B Endrin ke...	7.113	7.705	7433908	13191613	3.581	1.998 #

AT
 12/23/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.