

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112922\
 Data File : PD073253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2022 10:07
 Operator : AR\AJ
 Sample : PIBLK418
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 29 20:35:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111722CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 17 06:19:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.699	3.458	44620443	1374.7E6	18.608	21.045
27)	SA Decachlor...	7.824	8.906	67486069	438.0E6	29.575	29.600
Target Compounds							
2)	A alpha-BHC	0.000	3.922	0	2618080	N.D.	<MDL
3)	MA gamma-BHC...	0.000	4.262	0	4711652	N.D.	<MDL
4)	MA Heptachlor	0.000	4.819	0	1950396	N.D.	<MDL
5)	MB Aldrin	0.000	5.159	0	342119	N.D.	<MDL
6)	B beta-BHC	0.000	4.505f	0	24041798	N.D.	0.900
7)	B delta-BHC	0.000	4.722	0	3265714	N.D.	<MDL
8)	B Heptachlo...	0.000	5.591	0	66673531	N.D.	1.972
9)	A Endosulfan I	0.000	5.985	0	1126769	N.D.	<MDL
11)	B cis-Chlor...	0.000	5.937	0	1556261	N.D.	<MDL
13)	MA Dieldrin	0.000	6.272	0	13692304	N.D.	0.454
15)	B Endosulfa...	5.815f	6.724	92857	1039834	<MDL	<MDL #
16)	A 4,4'-DDD	5.654f	6.632	215463	673250	<MDL	<MDL #
18)	B Endrin al...	0.000	6.855	0	248363	N.D.	<MDL
19)	B Endosulfa...	0.000	7.099	0	3032865	N.D.	0.146
20)	A Methoxychlor	0.000	7.349f	0	3894407	N.D.	0.417
21)	B Endrin ke...	6.732f	7.597	88777	2518769	<MDL	0.113 #
22)	Toxaphene-1	0.000	5.766	0	10536401	N.D.	89.665
24)	Toxaphene-3	0.000	6.980	0	1466464	N.D.	2.921
25)	Toxaphene-4	0.000	7.099f	0	3032865	N.D.	6.792
26)	Toxaphene-5	0.000	7.523f	0	316350	N.D.	1.158


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112922\
Data File : PD073253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2022 10:07
Operator : AR\AJ
Sample : PIBLK418
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 29 20:35:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111722CLP.M
Quant Title : GC Extractables
QLast Update : Thu Nov 17 06:19:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.699 min
Delta R.T.: 0.000 min
Response: 44620443
Conc: 18.61 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: -0.001 min
Response: 1374746165
Conc: 21.04 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 3.834 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.505 min
Delta R.T.: 0.032 min
Response: 24041798
Conc: 0.90 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.649 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.591 min
Delta R.T.: -0.004 min
Response: 66673531
Conc: 1.97 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T.: 5.283 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 6.272 min
Delta R.T.: 0.016 min
Response: 13692304
Conc: 0.45 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T. : 6.264 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.099 min
Delta R.T.: 0.017 min
Response: 3032865
Conc: 0.15 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.539 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.349 min
Delta R.T.: -0.060 min
Response: 3894407
Conc: 0.42 ng/ml



#21 Endrin ketone


R.T.: 6.732 min
Delta R.T.: -0.042 min
Response: 88777
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#21 Endrin ketone

R.T.: 7.597 min
Delta R.T.: 0.026 min
Response: 2518769
Conc: 0.11 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T. : 5.246 min
Response: 0
Conc: N.D.



#22 Toxaphene-1


R.T.: 5.766 min
Delta R.T.: -0.002 min
Response: 10536401
Conc: 89.67 ng/ml



#23 Toxaphene-2


R.T.: 5.591 min
Delta R.T.: -0.014 min
Response: -158494
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#23 Toxaphene-2

R.T.: 6.158 min
Delta R.T.: 0.014 min
Response: 1321233
Conc: 6.17 ng/ml




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 5.890 min
Response: 0
Conc: N.D.



#24 Toxaphene-3


R.T.: 6.980 min
Delta R.T.: 0.013 min
Response: 1466464
Conc: 2.92 ng/ml



#25 Toxaphene-4


R.T.: 0.000 min
Exp R.T.: 6.649 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#25 Toxaphene-4

R.T.: 7.099 min
Delta R.T.: 0.042 min
Response: 3032865
Conc: 6.79 ng/ml




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T.: 6.960 min
Response: 0
Conc: N.D.



#26 Toxaphene-5

R.T.: 7.523 min
Delta R.T.: 0.038 min
Response: 316350
Conc: 1.16 ng/ml



#27 Decachlorobiphenyl


R.T.: 7.824 min

Delta R.T.: -0.005 min

Response: 67486069

Conc: 29.57 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 8.906 min

Delta R.T.: -0.007 min

Response: 438039888

Conc: 29.60 ng/ml