

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX113021.M
 Title : GC Extractables
 Last Update : Tue Nov 30 11:37:37 2021
 Response Via : Initial Calibration

Calibration Files

500 =PD067141.D 1000 =PD067139.D 750 =PD067140.D
 250 =PD067142.D 100 =PD067143.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.672	2.608	2.585	2.602	2.844	2.662	E6	4.01
2)	Toxaphene-1	2.221	2.226	2.229	2.395	2.692	2.353	E4	8.66
3)	Toxaphene-2	3.727	3.419	3.359	3.879	4.444	3.766	E4	11.58
4)	Toxaphene-3	5.335	5.127	5.138	5.342	6.233	5.435	E4	8.43
5)	Toxaphene-4	5.409	5.297	5.273	5.718	6.922	5.724	E4	12.11
6)	Toxaphene-5	5.496	5.223	5.224	5.567	6.032	5.508	E4	6.03
7) SA	Decachlorobiphenyl	1.371	1.280	1.281	1.433	1.726	1.418	E6	12.96

Signal #2 Calibration Files

500 =PD067141.D 1000 =PD067139.D 750 =PD067140.D
 250 =PD067142.D 100 =PD067143.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	0.877	0.807	0.828	0.908	1.061	0.896	E7	11.22
2)	Toxaphene-1	4.452	4.398	4.641	4.436	4.691	4.524	E4	2.93
3)	Toxaphene-2	7.456	6.771	6.866	7.405	8.187	7.337	E4	7.72
4)	Toxaphene-3	1.391	1.315	1.300	1.408	1.509	1.385	E5	6.05
5)	Toxaphene-4	7.960	7.539	7.536	7.832	7.898	7.753	E4	2.61
6)	Toxaphene-5	1.178	1.100	1.064	1.176	1.298	1.164	E5	7.73
7) SA	Decachlorobiphenyl	6.103	5.590	5.664	6.361	7.456	6.235	E6	12.07

(#) = Out of Range