

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD120219CLP.M
 Title : GC Extractables
 Last Update : Tue Dec 03 06:52:11 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD056202.D 1 =PD056200.D 2 =PD056201.D
 4 =PD056203.D 5 =PD056204.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	0.999	1.027	1.007	0.975	1.008	1.003	E6 1.85
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	8.820	9.202	8.621	8.863	9.543	9.010	E3 4.04
23)	Toxaphene-2	9.537	9.938	9.469	9.751	9.959	9.731	E3 2.31
24)	Toxaphene-3	8.346	8.358	7.899	8.128	8.644	8.275	E3 3.37
25)	Toxaphene-4	2.457	2.400	2.365	2.535	2.631	2.478	E4 4.31
26)	Toxaphene-5	2.978	2.932	2.896	3.055	3.228	3.018	E4 4.37
27) SA	Decachlorobiphenyl	1.062	1.144	1.087	1.027	1.042	1.073	E6 4.28

Signal #2 Calibration Files

3 =PD056202.D 1 =PD056200.D 2 =PD056201.D
 4 =PD056203.D 5 =PD056204.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.455	1.484	1.453	1.418	1.449	1.452	E6 1.63
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	1.118	1.145	1.103	1.096	1.151	1.123	E4 2.20
23)	Toxaphene-2	1.411	1.372	1.429	1.360	1.327	1.380	E4 2.96
24)	Toxaphene-3	2.005	1.963	1.946	1.931	1.991	1.967	E4 1.57

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD120219CLP.M
 Title : GC Extractables
 Last Update : Tue Dec 03 06:52:11 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD056202.D 1 =PD056200.D 2 =PD056201.D
 4 =PD056203.D 5 =PD056204.D

	Compound	3	1	2	4	5	Avg		%RSD
25)	Toxaphene-4	5.538	5.386	5.445	5.575	5.665	5.522	E4	1.98
26)	Toxaphene-5	2.513	2.540	2.549	2.585	2.592	2.556	E4	1.27
27) SA	Decachlorobiphenyl	1.402	1.521	1.451	1.329	1.315	1.404	E6	6.10

(#) = Out of Range