

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120219\
 Data File : PD056209.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2019 16:19
 Operator : SG\AJ
 Sample : INDA301
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA301

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 08:40:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:39:41 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.288	3.963	15918151	23107455	20.000	20.000
27)	SA Decachlor...	7.965	9.002	34986770	47024089	40.000	40.000
Target Compounds							
2)	A alpha-BHC	3.715	4.349	22447161	31852435	20.000	20.000
3)	MA gamma-BHC...	3.991	4.632	21264753	31021183	20.000	20.000
4)	MA Heptachlor	4.276	5.139	19266416	31320012	20.000	20.000
9)	A Endosulfan I	5.284	6.182	17946956	28106309	20.000	20.000
13)	MA Dieldrin	5.524	6.438	39291084	61033038	40.000	40.000
14)	MA Endrin	5.778	6.655	32423979	49259434	40.000	40.000
16)	A 4,4'-DDD	5.912	6.778	27651114	46068106	40.000	40.000
17)	MA 4,4'-DDT	6.147	7.076	27555497	46068036	40.000	40.000
20)	A Methoxychlor	6.696	7.533	66826148	119.3E6	200.000	200.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120219\
Data File : PD056209.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2019 16:19
Operator : SG\AJ
Sample : INDA301
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA301

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 08:40:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:39:41 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.288 min
Delta R.T.: 0.000 min
Response: 15918151
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301




#1 Tetrachloro-m-xylene

R.T.: 3.963 min
Delta R.T.: 0.000 min
Response: 23107455
Conc: 20.00 ng/ml




#2 alpha-BHC

R.T.: 3.715 min
Delta R.T.: 0.000 min
Response: 22447161
Conc: 20.00 ng/ml



#2 alpha-BHC


R.T.: 4.349 min
Delta R.T.: 0.000 min
Response: 31852435
Conc: 20.00 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.991 min
Delta R.T.: 0.000 min
Response: 21264753
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301




#3 gamma-BHC (Lindane)

R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 31021183
Conc: 20.00 ng/ml




#4 Heptachlor

R.T.: 4.276 min
Delta R.T.: 0.000 min
Response: 19266416
Conc: 20.00 ng/ml



#4 Heptachlor


R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 31320012
Conc: 20.00 ng/ml



#9 Endosulfan I


R.T.: 5.284 min
Delta R.T.: 0.000 min
Response: 17946956
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301




#9 Endosulfan I

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 28106309
Conc: 20.00 ng/ml




#13 Dieldrin

R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 39291084
Conc: 40.00 ng/ml



#13 Dieldrin


R.T.: 6.438 min
Delta R.T.: 0.000 min
Response: 61033038
Conc: 40.00 ng/ml



#14 Endrin


R.T.: 5.778 min
Delta R.T.: 0.000 min
Response: 32423979
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301




#14 Endrin

R.T.: 6.655 min
Delta R.T.: 0.000 min
Response: 49259434
Conc: 40.00 ng/ml




#16 4,4'-DDD

R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 27651114
Conc: 40.00 ng/ml



#16 4,4'-DDD


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 46068106
Conc: 40.00 ng/ml



#17 4,4'-DDT


R.T.: 6.147 min
Delta R.T.: 0.000 min
Response: 27555497
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301




#17 4,4'-DDT

R.T.: 7.076 min
Delta R.T.: 0.000 min
Response: 46068036
Conc: 40.00 ng/ml




#20 Methoxychlor

R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 66826148
Conc: 200.00 ng/ml



#20 Methoxychlor


R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 119297977
Conc: 200.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.965 min
Delta R.T.: 0.000 min
Response: 34986770
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDA301



#27 Decachlorobiphenyl

R.T.: 9.002 min
Delta R.T.: 0.000 min
Response: 47024089
Conc: 40.00 ng/ml