

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120219\
 Data File : PD056213.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2019 17:15
 Operator : SG\AJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 08:41:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:39:41 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.288	3.963	69159856	99674812	86.894	86.271
27)	SA Decachlor...	7.965	9.001	143.4E6	185.6E6	163.937	157.863
Target Compounds							
2)	A alpha-BHC	3.715	4.349	106.0E6	151.5E6	94.409	95.138
3)	MA gamma-BHC...	3.991	4.632	97019458	142.6E6	91.249	91.930
4)	MA Heptachlor	4.276	5.139	86865637	139.1E6	90.173	88.856
9)	A Endosulfan I	5.284	6.182	79075207	122.6E6	88.121	87.272
13)	MA Dieldrin	5.524	6.438	179.9E6	275.4E6	183.197	180.505
14)	MA Endrin	5.777	6.655	143.1E6	219.2E6	176.596	177.984
16)	A 4,4'-DDD	5.911	6.777	125.0E6	208.3E6	180.770	180.869
17)	MA 4,4'-DDT	6.147	7.075	134.7E6	213.2E6	195.538	185.137
20)	A Methoxychlor	6.695	7.532	277.0E6	486.6E6	829.107	815.761

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120219\
Data File : PD056213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2019 17:15
Operator : SG\AJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 08:41:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:39:41 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

