

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120220\
 Data File : PD060554.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2020 21:26
 Operator : DD\AJ
 Sample : INDA330
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA330

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 05:05:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112720CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 28 01:50:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.502	4.023	21777958	73371759	19.125	19.957
27)	SA Decachlor...	8.292	9.107	53515571	111.4E6	39.518	38.423
Target Compounds							
2)	A alpha-BHC	3.944	4.413	31020067	108.5E6	18.913	19.893
3)	MA gamma-BHC...	4.230	4.699	29342788	98503399	18.845	19.897
4)	MA Heptachlor	4.536	5.210	30041122	92921283	19.241	19.179
9)	A Endosulfan I	5.574	6.261	26999658	78537153	19.607	20.077
13)	MA Dieldrin	5.819	6.519	58785952	170.4E6	38.623	40.167
14)	MA Endrin	6.078	6.737	41546049	115.7E6	33.377	33.350
16)	A 4,4'-DDD	6.202	6.856	49768875	140.2E6	40.913	42.329
17)	MA 4,4'-DDT	6.445	7.156	42818234	113.4E6	35.053	34.539
20)	A Methoxychlor	6.993	7.612	111.8E6	286.0E6	187.125	174.443

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120220\
Data File : PD060554.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2020 21:26
Operator : DD\AJ
Sample : INDA330
Misc :
ALS Vial : 29 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 05:05:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112720CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 28 01:50:00 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

