

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120318\
 Data File : PD050603.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Dec 2018 16:53
 Operator : AJ\SJ
 Sample : INDA101
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 18:34:40 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120318CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Dec 03 18:33:18 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	4.044	7751183	156.4E6	5.285	5.652
27)	SA Decachlor...	7.834	9.153	18368394	252.5E6	11.683	12.030
Target Compounds							
2)	A alpha-BHC	3.648	4.436	11753668	230.3E6	5.064	5.518
3)	MA gamma-BHC...	3.917	4.723	11762767	217.3E6	5.303	5.653
4)	MA Heptachlor	4.195	5.237	11579492	227.5E6	5.097	5.777
9)	A Endosulfan I	5.178	6.290	10350192	173.6E6	5.231	5.747
13)	MA Dieldrin	5.414	6.549	22512541	352.8E6	10.111	11.483
14)	MA Endrin	5.661	6.768	19102736	263.4E6	10.089	11.400
16)	A 4,4'-DDD	5.801	6.887	16471430	268.6E6	10.130	11.364
17)	MA 4,4'-DDT	6.033	7.187	17490752	242.4E6	9.996	10.745
20)	A Methoxychlor	6.580	7.645	50586285	619.0E6	53.581	58.255

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120318\
Data File : PD050603.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Dec 2018 16:53
Operator : AJ\SJ
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 18:34:40 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120318CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 03 18:33:18 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

