

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120318\
 Data File : PD050607.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Dec 2018 17:49
 Operator : AJ\SJ
 Sample : INDA301
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA301

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 18:35:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120318CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Dec 03 18:33:18 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.228	4.044	29331011	553.4E6	20.000	20.000
27)	SA Decachlor...	7.834	9.153	62890352	839.6E6	40.000	40.000
Target Compounds							
2)	A alpha-BHC	3.648	4.436	46417383	834.8E6	20.000	20.000
3)	MA gamma-BHC...	3.917	4.723	44359168	768.9E6	20.000	20.000
4)	MA Heptachlor	4.195	5.237	45437032	787.6E6	20.000	20.000
9)	A Endosulfan I	5.178	6.290	39570550	604.1E6	20.000	20.000
13)	MA Dieldrin	5.414	6.549	89065665	1228.8E6	40.000	40.000
14)	MA Endrin	5.661	6.767	75734923	924.2E6	40.000	40.000
16)	A 4,4'-DDD	5.801	6.887	65039531	945.3E6	40.000	40.000
17)	MA 4,4'-DDT	6.033	7.187	69992282	902.5E6	40.000	40.000
20)	A Methoxychlor	6.580	7.645	188.8E6	2125.2E6	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120318\
Data File : PD050607.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Dec 2018 17:49
Operator : AJ\SJ
Sample : INDA301
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 18:35:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120318CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 03 18:33:18 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSample :
INDA301