

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120623\
 Data File : PD079955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 13:52
 Operator : AR\AJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 14:08:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120623.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 06 14:08:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.801	7050419	13356399	4.886	4.736
28)	SA Decachlor...	9.114	7.944	10084764	14108121	5.234	5.182
Target Compounds							
23)	Chlordane-1	4.734	3.801	3686657	4508711	49.498	46.835
24)	Chlordane-2	5.265	4.377	4207216	6057993	55.041	53.000
25)	Chlordane-3	5.977	5.007	14499194	14261694	48.024	45.269
26)	Chlordane-4	6.059	5.071	17940201	14172976	49.538	48.171
27)	Chlordane-5	6.910	5.966	3405408	5275046	50.338	46.432



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120623\
Data File : PD079955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 13:52
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 14:08:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120623.M
Quant Title : GC Extractables
QLast Update : Wed Dec 06 14:08:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.567 min
Delta R.T.: 0.000 min
Response: 7050419
Conc: 4.89 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#1 Tetrachloro-m-xylene

R.T.: 2.801 min
Delta R.T.: 0.000 min
Response: 13356399
Conc: 4.74 ng/ml




#23 Chlordane-1

R.T.: 4.734 min
Delta R.T.: 0.000 min
Response: 3686657
Conc: 49.50 ng/ml



#23 Chlordane-1


R.T.: 3.801 min
Delta R.T.: 0.000 min
Response: 4508711
Conc: 46.84 ng/ml



#24 Chlordane-2


R.T.: 5.265 min
Delta R.T.: 0.000 min
Response: 4207216
Conc: 55.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#24 Chlordane-2

R.T.: 4.377 min
Delta R.T.: 0.000 min
Response: 6057993
Conc: 53.00 ng/ml




#25 Chlordane-3

R.T.: 5.977 min
Delta R.T.: 0.000 min
Response: 14499194
Conc: 48.02 ng/ml



#25 Chlordane-3


R.T.: 5.007 min
Delta R.T.: 0.000 min
Response: 14261694
Conc: 45.27 ng/ml



#26 Chlordane-4


R.T.: 6.059 min
Delta R.T.: 0.000 min
Response: 17940201
Conc: 49.54 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050




#26 Chlordane-4

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 14172976
Conc: 48.17 ng/ml




#27 Chlordane-5

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 3405408
Conc: 50.34 ng/ml



#27 Chlordane-5


R.T.: 5.966 min
Delta R.T.: 0.000 min
Response: 5275046
Conc: 46.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.114 min
Delta R.T.: 0.000 min
Response: 10084764
Conc: 5.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC050



#28 Decachlorobiphenyl

R.T.: 7.944 min
Delta R.T.: 0.000 min
Response: 14108121
Conc: 5.18 ng/ml