

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120624\
 Data File : PD087050.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 13:18
 Operator : AR\AJ
 Sample : PB165422BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB165422BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 22:19:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112724.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 15:47:26 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.558	2.884	37232404	234.6E6	18.618	20.493
28)	SA Decachlor...	9.090	8.087	60097738	278.5E6	18.893	19.986
Target Compounds							
2)	A alpha-BHC	4.008	3.397	204.5E6	896.0E6	50.822	50.848
3)	MA gamma-BHC...	4.339	3.734	196.5E6	836.0E6	49.272	49.777
4)	MA Heptachlor	4.939	4.089	200.3E6	820.6E6	51.022	50.483
5)	MB Aldrin	5.282	4.376	196.4E6	817.9E6	48.189	48.926
6)	B beta-BHC	4.522	4.029	77001146	367.3E6	48.821	51.288
7)	B delta-BHC	4.771	4.267	195.5E6	824.0E6	47.951	48.464
8)	B Heptachlo...	5.702	4.880	177.4E6	747.6E6	49.088	49.237
9)	A Endosulfan I	6.085	5.255	167.8E6	711.3E6	50.003	50.102
10)	B gamma-Chl...	5.956	5.134	178.7E6	795.1E6	50.593	51.111
11)	B alpha-Chl...	6.038	5.199	179.1E6	768.9E6	50.121	50.316
12)	B 4,4'-DDE	6.207	5.385	163.6E6	761.5E6	50.291	50.517
13)	MA Dieldrin	6.358	5.522	181.1E6	777.4E6	50.130	49.925
14)	MA Endrin	6.586	5.798	140.0E6	657.3E6	46.256	46.987
15)	B Endosulfa...	6.797	6.089	151.3E6	684.7E6	50.612	49.745
16)	A 4,4'-DDD	6.715	5.939	129.9E6	635.2E6	49.985	50.537
17)	MA 4,4'-DDT	7.032	6.194	134.7E6	650.7E6	47.697	48.798
18)	B Endrin al...	6.925	6.268	117.1E6	529.1E6	47.919	47.689
19)	B Endosulfa...	7.160	6.492	141.6E6	655.5E6	48.669	48.452
20)	A Methoxychlor	7.505	6.766	73158053	326.6E6	45.744	45.526
21)	B Endrin ke...	7.642	7.001	158.8E6	724.8E6	48.933	48.093
22)	Mirex	8.127	7.198	109.1E6	537.7E6	43.716	43.607

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120624\
Data File : PD087050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 13:18
Operator : AR\AJ
Sample : PB165422BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB165422BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 22:19:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112724.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 15:47:26 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

