

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120624\
 Data File : PD087057.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 14:56
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 22:22:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112724.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 27 15:47:26 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.557	2.884	103.3E6	586.6E6	51.644	51.246
28)	SA Decachlor...	9.087	8.087	146.0E6	636.1E6	45.884	45.652
Target Compounds							
2)	A alpha-BHC	4.007	3.398	216.8E6	916.9E6	53.865	52.035
3)	MA gamma-BHC...	4.338	3.734	212.5E6	863.2E6	53.284	51.398
4)	MA Heptachlor	4.938	4.090	203.8E6	812.6E6	51.906	49.989
5)	MB Aldrin	5.281	4.376	212.2E6	862.7E6	52.059	51.604
6)	B beta-BHC	4.521	4.030	80498874	362.5E6	51.039	50.614
7)	B delta-BHC	4.770	4.267	215.9E6	877.1E6	52.970	51.590
8)	B Heptachlo...	5.701	4.881	185.7E6	760.4E6	51.382	50.082
9)	A Endosulfan I	6.084	5.256	170.0E6	713.3E6	50.648	50.245
10)	B gamma-Chl...	5.955	5.134	180.0E6	785.7E6	50.962	50.509
11)	B alpha-Chl...	6.037	5.199	180.8E6	767.8E6	50.592	50.242
12)	B 4,4'-DDE	6.205	5.385	164.1E6	779.8E6	50.449	51.730
13)	MA Dieldrin	6.357	5.522	182.4E6	782.9E6	50.481	50.277
14)	MA Endrin	6.584	5.798	148.8E6	683.2E6	49.159	48.839
15)	B Endosulfa...	6.795	6.090	150.5E6	675.6E6	50.355	49.087
16)	A 4,4'-DDD	6.715	5.940	132.6E6	644.8E6	51.011	51.300
17)	MA 4,4'-DDT	7.031	6.194	131.5E6	622.8E6	46.537	46.705
18)	B Endrin al...	6.924	6.268	118.6E6	538.4E6	48.502	48.534
19)	B Endosulfa...	7.159	6.492	141.5E6	657.0E6	48.624	48.559
20)	A Methoxychlor	7.504	6.765	71957142	334.4E6	44.993	46.618
21)	B Endrin ke...	7.640	7.001	157.7E6	723.3E6	48.579	47.993
22)	Mirex	8.125	7.197	116.5E6	571.8E6	46.688	46.375

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD120624\
Data File : PD087057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 14:56
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 22:22:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112724.M
Quant Title : GC Extractables
QLast Update : Wed Nov 27 15:47:26 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

