

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX120825.M
 Title : GC Extractables
 Last Update : Tue Dec 09 04:12:30 2025
 Response Via : Initial Calibration

Calibration Files

500 =PD091493.D 1000 =PD091491.D 750 =PD091492.D
 250 =PD091494.D 100 =PD091495.D

Compound			500	1000	750	250	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		3.280	3.297	3.237	3.436	3.294	3.309	E6	2.27
2)	Toxaphene-1		3.781	3.563	3.678	4.230	4.668	3.984	E4	11.50
3)	Toxaphene-2		5.592	5.596	5.470	5.925	5.687	5.654	E4	3.01
4)	Toxaphene-3		1.035	1.064	1.032	1.079	1.025	1.047	E5	2.22
5)	Toxaphene-4		1.337	1.363	1.339	1.357	1.235	1.326	E5	3.95
6)	Toxaphene-5		7.687	7.798	7.562	8.046	7.413	7.701	E4	3.12
7) SA	Decachlorobiphenyl		4.929	4.775	4.755	5.368	5.198	5.005	E6	5.38

Signal #2 Calibration Files

500 =PD091493.D 1000 =PD091491.D 750 =PD091492.D
 250 =PD091494.D 100 =PD091495.D

Compound			500	1000	750	250	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		3.740	3.682	3.650	4.001	3.873	3.789	E7	3.85
2)	Toxaphene-1		3.181	3.078	3.072	3.521	3.316	3.234	E5	5.84
3)	Toxaphene-2		2.189	2.102	2.075	2.362	2.158	2.177	E5	5.19
4)	Toxaphene-3		1.072	1.071	1.054	1.139	1.027	1.073	E6	3.84
5)	Toxaphene-4		6.907	6.706	6.640	7.275	6.812	6.868	E5	3.63
6)	Toxaphene-5		5.117	5.164	5.011	5.289	4.363	4.989	E5	7.30
7) SA	Decachlorobiphenyl		4.213	4.262	4.186	4.497	4.386	4.309	E7	3.02

(#) = Out of Range