

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121024\
 Data File : PD087086.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2024 14:44
 Operator : AR\AJ
 Sample : INDA333
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3171

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 00:40:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111424CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 14 03:38:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.561	2.885	36303660	222.7E6	24.174	21.419
2)	SA Decachlor...	9.090	8.087	108.1E6	508.9E6	48.834	42.896
Target Compounds							
2)	A alpha-BHC	4.010	3.399	69804750	349.2E6	24.283	21.532
3)	MA gamma-BHC...	4.341	3.735	68376018	322.3E6	24.133	21.235
4)	MA Heptachlor	4.941	4.090	69691804	321.8E6	24.233	20.502
9)	A Endosulfan I	6.087	5.256	59538583	274.4E6	24.609	20.843
13)	MA Dieldrin	6.359	5.522	129.5E6	583.4E6	48.287	41.669
14)	MA Endrin	6.587	5.799	100.2E6	506.2E6	44.715	39.441
16)	A 4,4'-DDD	6.717	5.940	93879949	480.7E6	48.278	42.839
17)	MA 4,4'-DDT	7.033	6.195	88823953	458.6E6	43.732	38.434
20)	A Methoxychlor	7.505	6.767	248.7E6	1026.5E6	218.786	186.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121024\
Data File : PD087086.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2024 14:44
Operator : AR\AJ
Sample : INDA333
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 00:40:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111424CLP.M
Quant Title : GC Extractables
QLast Update : Thu Nov 14 03:38:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

