

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD121025\  
 Data File : PD091531.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 10 Dec 2025 12:33  
 Operator : AR\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Dec 11 01:53:35 2025  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD120825.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Dec 09 04:29:17 2025  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|---------|--------|--------|
| -----                       |                 |       |       |          |         |        |        |
| System Monitoring Compounds |                 |       |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 3.544 | 2.870 | 57276931 | 763.9E6 | 16.491 | 19.386 |
| 28)                         | SA Decachlor... | 9.069 | 8.059 | 90141581 | 893.6E6 | 17.342 | 21.579 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD121025\  
Data File : PD091531.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 10 Dec 2025 12:33  
Operator : AR\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_D  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Dec 11 01:53:35 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD120825.M  
Quant Title : GC Extractables  
QLast Update : Tue Dec 09 04:29:17 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min  
Delta R.T.: 0.004 min  
Response: 57276931  
Conc: 16.49 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.870 min  
Delta R.T.: 0.000 min  
Response: 763868901  
Conc: 19.39 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.069 min  
Delta R.T.: 0.009 min  
Response: 90141581  
Conc: 17.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min  
Delta R.T.: 0.004 min  
Response: 893585715  
Conc: 21.58 ng/ml