

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121119\
 Data File : PD056389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2019 16:56
 Operator : SG\AJ
 Sample : K6170-17
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0BS5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 01:17:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:51:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.287	3.961	19790320	29494278	23.896	24.532
27)	SA Decachlor...	7.959	8.995	34220409	45934966	37.898	38.179
Target Compounds							
3)	MA gamma-BHC...	0.000	4.594f	0	691147	N.D.	0.428
6)	B beta-BHC	4.208f	4.765f	6641523	489987	12.339	0.595 #
7)	B delta-BHC	0.000	4.990f	0	3751541	N.D.	2.252
8)	B Heptachlo...	0.000	5.814	0	2614453	N.D.	1.688
16)	A 4,4'-DDD	0.000	6.735f	0	2886929	N.D.	2.453
19)	B Endosulfa...	6.479f	7.237	4140704	5843097	4.767	4.370
21)	B Endrin ke...	6.960f	0.000	3434985	0	3.547	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121119\
Data File : PD056389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2019 16:56
Operator : SG\AJ
Sample : K6170-17
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0BS5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:17:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:51:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.287 min
Delta R.T.: 0.000 min
Response: 19790320
Conc: 23.90 ng/ml

Instrument :
ECD_D
ClientSampleId :
C0BS5




#1 Tetrachloro-m-xylene

R.T.: 3.961 min
Delta R.T.: -0.002 min
Response: 29494278
Conc: 24.53 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 3.991 min
Response: 0
Conc: N.D.



#3 gamma-BHC (Lindane)


R.T.: 4.594 min
Delta R.T.: -0.038 min
Response: 691147
Conc: 0.43 ng/ml



#6 beta-BHC


R.T.: 4.208 min
Delta R.T.: -0.032 min
Response: 6641523
Conc: 12.34 ng/ml

Instrument :
ECD_D
ClientSampleId :
C0BS5




#6 beta-BHC

R.T.: 4.765 min
Delta R.T.: -0.035 min
Response: 489987
Conc: 0.59 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.436 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.990 min
Delta R.T.: -0.027 min
Response: 3751541
Conc: 2.25 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 4.953 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
C0BS5




#8 Heptachlor epoxide

R.T.: 5.814 min
Delta R.T.: -0.012 min
Response: 2614453
Conc: 1.69 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 5.912 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.735 min
Delta R.T.: -0.042 min
Response: 2886929
Conc: 2.45 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.479 min
Delta R.T.: 0.048 min
Response: 4140704
Conc: 4.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
C0BS5




#19 Endosulfan Sulfate

R.T.: 7.237 min
Delta R.T.: 0.027 min
Response: 5843097
Conc: 4.37 ng/ml




#21 Endrin ketone

R.T.: 6.960 min
Delta R.T.: 0.044 min
Response: 3434985
Conc: 3.55 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 7.679 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 7.959 min
Delta R.T.: -0.006 min
Response: 34220409
Conc: 37.90 ng/ml

Instrument :
ECD_D
ClientSampleId :
C0BS5



#27 Decachlorobiphenyl

R.T.: 8.995 min
Delta R.T.: -0.007 min
Response: 45934966
Conc: 38.18 ng/ml