

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121123\
 Data File : PD080018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2023 11:11
 Operator : AR\AJ
 Sample : PEM052
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM063

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 12/12/2023
 Supervised By :Ankita Jodhani 12/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 21:39:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120523CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 05 17:09:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.568	2.801	27006192	42818733	18.880	18.681
27)	SA Decachlor...	9.119	7.946	36628860	45179713	19.590	17.463
Target Compounds							
2)	A alpha-BHC	4.025	3.304	21982278	33646578	9.115	9.171
3)	MA gamma-BHC...	4.359	3.634	21372032	31067268	9.237	9.080
6)	B beta-BHC	4.553	3.929	10144714	16179130	9.593	10.150
14)	MA Endrin	6.616	5.669	74980703	115.6E6	41.741	41.475
16)	A 4,4'-DDD	6.746	5.819	350137	859172	0.216m	0.350 #
17)	MA 4,4'-DDT	7.062	6.068	154.8E6	227.3E6	92.457	92.307
18)	B Endrin al...	6.963	6.142	1551169	3948016	1.037	1.865m#
20)	A Methoxychlor	7.536	6.642	206.7E6	292.3E6	227.400	220.517
21)	B Endrin ke...	7.685	6.870	4287279	6362138	2.139	2.149

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121123\
Data File : PD080018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2023 11:11
Operator : AR\AJ
Sample : PEM052
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 21:39:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120523CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 05 17:09:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Miza 12/12/2023
Supervised By: Ankita Jodhani 12/12/2023

Instrument :
ECD_D
Client Sampled :
PEM063