


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121219\
Data File : PD056420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 18:52
Operator : SG\AJ
Sample : K6202-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0BX3

Manual Integrations
APPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 00:54:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:51:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(3) gamma-BHC (Lindane) (MA)

0.000min 0.000 ng/ml

response 0

(3) gamma-BHC (Lindane) #2 (MA)

0.000min 0.000 ng/ml

response 0

QEdit

(+) = Expected Retention Time


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121219\
Data File : PD056420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 18:52
Operator : SG\AJ
Sample : K6202-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0BX3

Manual Integrations
APPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 00:54:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:51:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(3) gamma-BHC (Lindane) (MA)

3.996min 1.212 ng/ml m

response 1361158

(3) gamma-BHC (Lindane) #2 (MA)

4.631min 0.529 ng/ml m

response 854317

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121219\
 Data File : PD056420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2019 18:52
 Operator : SG\AJ
 Sample : K6202-03
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0BX3

Manual Integrations
 APPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 00:54:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:51:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.287	3.960	18217206	27433933	21.996	22.819
27) SA Decachlor...	7.956	8.992	36118805	47267231	40.000	39.286
Target Compounds						
3) MA gamma-BHC...	3.996	4.631	1361158	854317	1.212m	0.529m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Ankita

12/13/2019 12:27:19 PM

3 AJ
 12/24/19