

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121220\
 Data File : PD060716.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2020 09:41
 Operator : DD\AJ
 Sample : PB133562BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS62

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 00:19:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 04 03:58:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.502	4.022	18304763	72196306	15.668	18.355
27)	SA Decachlor...	8.289	9.107	43148693	110.0E6	28.944	33.980
Target Compounds							
3)	MA gamma-BHC...	4.228	4.698	6983272	23932506	4.329	4.463
8)	B Heptachlo...	5.230	5.902	8348298	26799785	5.464	6.222
10)	B trans-Chl...	5.459	6.137	8536686	27567853	5.397	5.849
12)	B 4,4'-DDE	5.681	6.365	15627198	50777528	10.382	11.498
13)	MA Dieldrin	5.817	6.518	16065292	53679120	10.116	11.493
14)	MA Endrin	6.076	6.737	13753472	42534162	10.630	11.538
19)	B Endosulfa...	6.732	7.292	7385266	21503989	6.195	6.018

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121220\
Data File : PD060716.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2020 09:41
Operator : DD\AJ
Sample : PB133562BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 00:19:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 04 03:58:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

