

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121223\
 Data File : PD080032.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 09:48
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 13:58:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120623.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 06 14:11:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.568	2.800	30533835	49158854	20.259	19.556
28)	SA Decachlor...	9.113	7.943	39343390	52474994	19.437	18.615
Target Compounds							
2)	A alpha-BHC	4.025	3.303	25173497	38222764	10.022	9.996
3)	MA gamma-BHC...	4.358	3.633	23904158	36285429	10.089	9.985
6)	B beta-BHC	4.552	3.928	11416412	17934031	11.408	11.617
14)	MA Endrin	6.613	5.667	83844363	137.3E6	46.234	48.864
16)	A 4,4'-DDD	6.745	5.815	1087656	1969574	0.649	0.765
17)	MA 4,4'-DDT	7.059	6.065	170.3E6	263.6E6	95.719	103.740
18)	B Endrin al...	6.961	6.140	2302291	4810557	1.607	2.343 #
20)	A Methoxychlor	7.533	6.640	220.1E6	335.5E6	222.664	229.415
21)	B Endrin ke...	7.682	6.867	4905651	8123450	2.369	2.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121223\
Data File : PD080032.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 09:48
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 13:58:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120623.M
Quant Title : GC Extractables
QLast Update : Wed Dec 06 14:11:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

