

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121319\
 Data File : PD056456.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2019 21:41
 Operator : SG\AJ
 Sample : K6240-23
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BFQT0

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 22:15:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:51:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.287	3.961	20399381	28098049	24.631	23.371
27)	SA Decachlor...	7.960	8.997	39551433	52672458	43.801	43.779
Target Compounds							
5)	MB Aldrin	0.000	5.486f	0	5961755	N.D.	3.628
6)	B beta-BHC	0.000	4.798	0	889061	N.D.	1.079
7)	B delta-BHC	0.000	5.060f	0	3802997	N.D.	2.283
8)	B Heptachlo...	0.000	5.794	0	14645702	N.D.	9.453
15)	B Endosulfa...	0.000	6.872	0	766241	N.D.	0.523
16)	A 4,4'-DDD	0.000	6.734f	0	468005	N.D.	0.398
19)	B Endosulfa...	0.000	7.182	0	677933	N.D.	0.507


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121319\
Data File : PD056456.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2019 21:41
Operator : SG\AJ
Sample : K6240-23
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BFQT0

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 22:15:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:51:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.287 min
Delta R.T.: 0.000 min
Response: 20399381
Conc: 24.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
BFQT0




#1 Tetrachloro-m-xylene

R.T.: 3.961 min
Delta R.T.: -0.001 min
Response: 28098049
Conc: 23.37 ng/ml




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 4.516 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 5.486 min
Delta R.T.: 0.044 min
Response: 5961755
Conc: 3.63 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.240 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BFQT0




#6 beta-BHC

R.T.: 4.798 min
Delta R.T.: -0.002 min
Response: 889061
Conc: 1.08 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.436 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.060 min
Delta R.T.: 0.043 min
Response: 3802997
Conc: 2.28 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.953 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BFQT0




#8 Heptachlor epoxide

R.T.: 5.794 min
Delta R.T.: -0.032 min
Response: 14645702
Conc: 9.45 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 6.051 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.872 min
Delta R.T.: 0.008 min
Response: 766241
Conc: 0.52 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 5.912 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BFQT0




#16 4,4'-DDD

R.T.: 6.734 min
Delta R.T.: -0.043 min
Response: 468005
Conc: 0.40 ng/ml




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T.: 6.430 min
Response: 0
Conc: N.D.



#19 Endosulfan Sulfate


R.T.: 7.182 min
Delta R.T.: -0.028 min
Response: 677933
Conc: 0.51 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.960 min
Delta R.T.: -0.005 min
Response: 39551433
Conc: 43.80 ng/ml

Instrument :
ECD_D
ClientSampleId :
BFQT0



#27 Decachlorobiphenyl

R.T.: 8.997 min
Delta R.T.: -0.005 min
Response: 52672458
Conc: 43.78 ng/ml