

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121321\
 Data File : PD067290.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2021 18:12
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 02:29:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD113021.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 17:33:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.435	3.966	99290652	439.7E6	48.052	60.574 #
28)	SA Decachlor...	8.158	9.026	48692029	250.7E6	42.976	49.864
Target Compounds							
2)	A alpha-BHC	3.872	4.359	167.1E6	619.0E6	49.755	59.477
3)	MA gamma-BHC...	4.154	4.644	153.2E6	539.5E6	47.687	56.611
4)	MA Heptachlor	4.447	5.144	152.1E6	447.8E6	46.748	49.752
5)	MB Aldrin	4.691	5.448	155.4E6	418.5E6	48.788	49.782
6)	B beta-BHC	4.402	4.822	63919096	233.3E6	46.827	52.919
7)	B delta-BHC	4.604	5.037	149.4E6	483.9E6	49.961	52.731
8)	B Heptachlo...	5.132	5.837	138.4E6	361.1E6	48.031	51.499
9)	A Endosulfan I	5.468	6.194	125.0E6	347.2E6	47.953	49.905
10)	B gamma-Chl...	5.356	6.071	138.9E6	365.4E6	47.774	49.648
11)	B alpha-Chl...	5.414	6.143	134.7E6	360.4E6	47.406	49.472
12)	B 4,4'-DDE	5.578	6.303	127.0E6	345.0E6	47.342	49.862
13)	MA Dieldrin	5.710	6.453	134.4E6	360.3E6	48.083	49.134
14)	MA Endrin	5.967	6.671	115.2E6	289.6E6	50.463	51.572
15)	B Endosulfa...	6.240	6.887	101.9E6	287.1E6	45.481	47.829
16)	A 4,4'-DDD	6.092	6.798	97930426	275.5E6	46.599	47.102
17)	MA 4,4'-DDT	6.331	7.096	101.9E6	283.7E6	48.141	47.528
18)	B Endrin al...	6.409	7.011	74252525	217.8E6	44.401	46.445
19)	B Endosulfa...	6.621	7.234	91751195	280.0E6	45.239	46.630
20)	A Methoxychlor	6.878	7.558	48372207	147.2E6	43.609	45.706
21)	B Endrin ke...	7.112	7.707	100.1E6	324.2E6	44.168	44.821
22)	Mirex	7.292	8.155	69321437	233.0E6	43.214	47.573

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121321\
Data File : PD067290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2021 18:12
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 02:29:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD113021.M
Quant Title : GC Extractables
QLast Update : Tue Nov 30 17:33:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

