

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121520\
 Data File : PD060730.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2020 14:02
 Operator : DD\AJ
 Sample : INDB341
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB341

Manual Integrations
 APPROVED

Ankita
 12/16/2020 10:58:33 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 01:16:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 04 03:58:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.502	4.021	23139676	92293354	19.806	23.464
27)	SA Decachlor...	8.287	9.105	56453497	146.0E6	37.869	45.107
Target Compounds							
5)	MB Aldrin	4.785	5.515	32718713	117.9E6	20.342	23.672
6)	B beta-BHC	4.478	4.865	15091662	53578922	19.682	22.515
7)	B delta-BHC	4.685	5.084	29912329	117.1E6	18.679	24.557 #
8)	B Heptachlo...	5.230	5.902	30536112	107.0E6	19.986	24.835
10)	B trans-Chl...	5.458	6.136	32271804	111.0E6	20.403	23.554
11)	B cis-Chlor...	5.518	6.208	31797144	106.5E6	20.414	23.621m
12)	B 4,4'-DDE	5.681	6.363	61587483	206.1E6	40.916	46.671m
15)	B Endosulfa...	6.349	6.944	50618768	164.3E6	40.021	45.695
18)	B Endrin al...	6.518	7.068	42078398	135.0E6	39.204	44.452
19)	B Endosulfa...	6.732	7.291	46463888	154.3E6	38.977	43.187
21)	B Endrin ke...	7.227	7.761	57516162	167.0E6	39.731	43.982m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121520\
Data File : PD060730.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2020 14:02
Operator : DD\AJ
Sample : INDB341
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 01:16:00 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 04 03:58:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
12/16/2020 10:58:33 AM

Instrument :
ECD_D
ClientSampled :
INDB341