

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121521\
 Data File : PD067371.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2021 01:18
 Operator : AR\AJ
 Sample : PIBLK180
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 04:31:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 05:28:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	3.965	48688128	239.8E6	22.298	29.937 #
27)	SA Decachlor...	8.156	9.022	45796325	241.9E6	40.337	47.694
Target Compounds							
4)	MA Heptachlor	4.408f	0.000	1576497	0	0.492	N.D. #
6)	B beta-BHC	4.408	4.840	1576497	1466529	1.184	0.318 #
7)	B delta-BHC	0.000	4.999f	0	17562491	N.D.	1.984
8)	B Heptachlo...	5.203f	5.833	3841666	38237089	1.391	5.456 #
10)	B trans-Chl...	0.000	6.023f	0	620083	N.D.	<MDL
12)	B 4,4'-DDE	0.000	6.348f	0	1039416	N.D.	0.155
16)	A 4,4'-DDD	0.000	6.846f	0	1314085	N.D.	0.225
22)	Toxaphene-1	0.000	6.348	0	1039416	N.D.	31.610

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121521\
Data File : PD067371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2021 01:18
Operator : AR\AJ
Sample : PIBLK180
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 04:31:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
Quant Title : GC Extractables
QLast Update : Tue Nov 30 05:28:01 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

