

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121621\
 Data File : PD067375.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2021 10:43
 Operator : AR\AJ
 Sample : PIBLK181
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PIBLK185

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 01:14:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 30 05:28:01 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	3.965	44875879	226.4E6	20.552	28.259 #
27)	SA Decachlor...	8.166	9.029	43390549	237.1E6	38.218	46.767
Target Compounds							
3)	MA gamma-BHC...	0.000	4.603f	0	497050	N.D.	<MDL
4)	MA Heptachlor	4.416f	5.155	2406650	1089789	0.751	0.117 #
6)	B beta-BHC	4.416	0.000	2406650	0	1.808	N.D. #
7)	B delta-BHC	0.000	5.001f	0	21692003	N.D.	2.451
8)	B Heptachlo...	0.000	5.859	0	705100	N.D.	0.101
13)	MA Dieldrin	0.000	6.384f	0	1959913	N.D.	0.263
16)	A 4,4'-DDD	0.000	6.852f	0	1160677	N.D.	0.198
21)	B Endrin ke...	0.000	7.729	0	4437937	N.D.	0.672
22)	Toxaphene-1	0.000	6.384	0	1959913	N.D.	59.604
26)	Toxaphene-5	7.028	0.000	696258	0	17.729	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121621\
Data File : PD067375.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2021 10:43
Operator : AR\AJ
Sample : PIBLK181
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 01:14:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD112921CLP.M
Quant Title : GC Extractables
QLast Update : Tue Nov 30 05:28:01 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
PIBLK185