

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121622\
 Data File : PD073330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2022 10:58
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 16:59:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121622.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 16 15:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.696	3.458	48729777	2185.7E6	23.251	22.225
28)	SA Decachlor...	7.827	8.912	55996109	412.0E6	23.127	23.371
Target Compounds							
9)	A Endosulfan I	5.014	5.977	29196819	483.6E6	11.032	11.649
10)	B gamma-Chl...	4.894	5.846	30550274	491.9E6	10.707	11.005
12)	B 4,4'-DDE	5.146	6.096	59114100	959.8E6	21.831	22.711
13)	MA Dieldrin	5.281	6.255	63084564	961.5E6	22.355	22.888
19)	B Endosulfa...	6.261	7.080	49360951	543.9E6	20.767	19.879
20)	A Methoxychlor	6.537	7.408	138.0E6	1319.8E6	103.504	97.510
21)	B Endrin ke...	6.772	7.570	63309305	618.0E6	23.297	22.507



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121622\
Data File : PD073330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2022 10:58
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 16:59:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121622.M
Quant Title : GC Extractables
QLast Update : Fri Dec 16 15:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.696 min
Delta R.T.: 0.000 min
Response: 48729777
Conc: 23.25 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK




#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: -0.001 min
Response: 2185716027
Conc: 22.22 ng/ml




#9 Endosulfan I

R.T.: 5.014 min
Delta R.T.: 0.000 min
Response: 29196819
Conc: 11.03 ng/ml



#9 Endosulfan I


R.T.: 5.977 min
Delta R.T.: 0.000 min
Response: 483633028
Conc: 11.65 ng/ml



#10 gamma-Chlordane


R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 30550274
Conc: 10.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK




#10 gamma-Chlordane

R.T.: 5.846 min
Delta R.T.: 0.000 min
Response: 491899650
Conc: 11.00 ng/ml




#12 4,4'-DDE

R.T.: 5.146 min
Delta R.T.: 0.000 min
Response: 59114100
Conc: 21.83 ng/ml



#12 4,4'-DDE


R.T.: 6.096 min
Delta R.T.: 0.000 min
Response: 959775949
Conc: 22.71 ng/ml



#13 Dieldrin


R.T.: 5.281 min
Delta R.T.: 0.000 min
Response: 63084564
Conc: 22.35 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK




#13 Dieldrin

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 961516211
Conc: 22.89 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 49360951
Conc: 20.77 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.080 min
Delta R.T.: 0.000 min
Response: 543902183
Conc: 19.88 ng/ml



#20 Methoxychlor


R.T.: 6.537 min
Delta R.T.: 0.002 min
Response: 137951192
Conc: 103.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK




#20 Methoxychlor

R.T.: 7.408 min
Delta R.T.: 0.000 min
Response: 1319833298
Conc: 97.51 ng/ml




#21 Endrin ketone

R.T.: 6.772 min
Delta R.T.: 0.000 min
Response: 63309305
Conc: 23.30 ng/ml



#21 Endrin ketone


R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 617970349
Conc: 22.51 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.827 min
Delta R.T.: 0.001 min
Response: 55996109
Conc: 23.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK



#28 Decachlorobiphenyl

R.T.: 8.912 min
Delta R.T.: 0.000 min
Response: 411981520
Conc: 23.37 ng/ml