

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121720\
 Data File : PD060770.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2020 14:05
 Operator : DD\AJ
 Sample : L5088-02
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 03:04:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121620.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 17 03:53:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.501	4.022	22751575	98802744	20.926	21.024
28)	SA Decachlor...	8.290	9.107	30196888	83072709	21.375	22.235
Target Compounds							
4)	MA Heptachlor	0.000	5.209	0	654677	N.D.	0.101 #
6)	B beta-BHC	4.445f	4.862	5196900	2898312	7.346	1.044 #
7)	B delta-BHC	0.000	5.061f	0	14589604	N.D.	2.297 #
8)	B Heptachlo...	0.000	5.925f	0	1225275	N.D.	0.219 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121720\
Data File : PD060770.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2020 14:05
Operator : DD\AJ
Sample : L5088-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 03:04:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121620.M
Quant Title : GC Extractables
QLast Update : Thu Dec 17 03:53:08 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.501 min
Delta R.T.: 0.000 min
Response: 22751575
Conc: 20.93 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.022 min
Delta R.T.: 0.000 min
Response: 98802744
Conc: 21.02 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.533 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.209 min
Delta R.T.: 0.000 min
Response: 654677
Conc: 0.10 ng/ml



#6 beta-BHC


R.T.: 4.445 min
Delta R.T.: -0.032 min
Response: 5196900
Conc: 7.35 ng/ml

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.862 min
Delta R.T.: -0.003 min
Response: 2898312
Conc: 1.04 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.684 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.061 min
Delta R.T.: -0.023 min
Response: 14589604
Conc: 2.30 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.229 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.925 min
Delta R.T.: 0.023 min
Response: 1225275
Conc: 0.22 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.290 min
Delta R.T.: 0.003 min
Response: 30196888
Conc: 21.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.107 min
Delta R.T.: 0.001 min
Response: 83072709
Conc: 22.23 ng/ml