

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD121723CLP-TCLP.M
 Title : GC Extractables
 Last Update : Sat Dec 16 02:47:23 2023
 Response Via : Initial Calibration

Calibration Files

3 =PD080127.D 1 =PD080123.D 2 =PD080125.D
 4 =PD080129.D 5 =PD080131.D 6 =PD080133.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.371	1.438	1.437	1.358	1.370	1.372	1.391 E6	2.60
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin	2.139	2.237	2.200	2.185	2.235	2.276	2.212 E6	2.17
6) B beta-BHC	0.990	1.072	1.066	0.962	0.944	0.929	0.994 E6	6.21
7) B delta-BHC	2.185	2.203	2.223	2.292	2.420	2.532	2.309 E6	6.02
8) B Heptachlor ep...	2.035	2.277	2.150	2.023	2.030	2.037	2.092 E6	4.90
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane	1.963	2.119	2.057	1.971	1.994	2.027	2.022 E6	2.91
11) B cis-Chlordane	2.013	2.187	2.113	2.010	2.010	2.032	2.061 E6	3.56
12) B 4,4'-DDE	1.845	1.844	1.863	1.922	1.994	2.045	1.919 E6	4.40
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II	1.795	2.027	1.918	1.780	1.780	1.778	1.846 E6	5.62
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde	1.420	1.555	1.505	1.402	1.391	1.381	1.442 E6	4.92
19) B Endosulfan Su...	1.728	1.875	1.825	1.715	1.714	1.708	1.761 E6	4.03
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone	1.877	2.009	1.963	1.872	1.860	1.882	1.910 E6	3.19
22) Toxaphene-1							0.000	-1.00
23) Toxaphene-2							0.000	-1.00
24) Toxaphene-3							0.000	-1.00
25) Toxaphene-4							0.000	-1.00
26) Toxaphene-5							0.000	-1.00
27) SA Decachlorobip...	1.752	1.936	1.878	1.695	1.645	1.617	1.754 E6	7.33

Signal #2 Calibration Files

3 =PD080127.D 1 =PD080123.D 2 =PD080125.D
 4 =PD080129.D 5 =PD080131.D 6 =PD080133.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	2.367	2.547	2.480	2.361	2.361	2.353	2.411 E6	3.40
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin	3.400	3.464	3.474	3.472	3.547	3.557	3.486 E6	1.67
6) B beta-BHC	1.697	1.856	1.796	1.662	1.628	1.596	1.706 E6	5.91
7) B delta-BHC	3.700	3.710	3.724	3.837	3.972	4.043	3.831 E6	3.84
8) B Heptachlor ep...	3.243	3.372	3.360	3.256	3.288	3.262	3.297 E6	1.70
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane	3.146	3.255	3.242	3.189	3.253	3.279	3.227 E6	1.54
11) B cis-Chlordane	3.238	3.406	3.371	3.266	3.301	3.304	3.314 E6	1.91
12) B 4,4'-DDE	3.011	2.930	3.001	3.127	3.207	3.205	3.080 E6	3.77
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II	2.849	2.948	2.932	2.874	2.879	2.817	2.883 E6	1.71
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde	2.222	2.361	2.322	2.214	2.202	2.142	2.244 E6	3.64
19) B Endosulfan Su...	2.672	2.787	2.762	2.684	2.681	2.610	2.699 E6	2.39
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone	3.094	3.160	3.180	3.107	3.092	2.984	3.103 E6	2.21
22) Toxaphene-1							0.000	-1.00
23) Toxaphene-2							0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD121723CLP-TCLP.M
 Title : GC Extractables
 Last Update : Sat Dec 16 02:47:23 2023
 Response Via : Initial Calibration

Calibration Files

3 =PD080127.D 1 =PD080123.D 2 =PD080125.D
 4 =PD080129.D 5 =PD080131.D 6 =PD080133.D

	Compound	3	1	2	4	5	6	Avg	%RSD
24)	Toxaphene-3							0.000	-1.00
25)	Toxaphene-4							0.000	-1.00
26)	Toxaphene-5							0.000	-1.00
27) SA	Decachlorobip...	2.639	2.889	2.792	2.592	2.567	2.528	2.668 E6	5.32

(#) = Out of Range