

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121821\
 Data File : PD067424.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2021 19:39
 Operator : AR\AJ
 Sample : RESC007
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESC015

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 01:11:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121821CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 18 01:09:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	3.965	42606173	212.5E6	20.358	20.181
27)	SA Decachlor...	8.157	9.023	40571342	238.2E6	41.184	41.219
Target Compounds							
2)	A alpha-BHC	3.871	4.358	32312784	145.8E6	10.279	10.740
3)	MA gamma-BHC...	4.152	4.643	33520402	134.8E6	11.089	11.257
4)	MA Heptachlor	4.445	5.143	32624426	115.1E6	10.493	11.098
5)	MB Aldrin	4.689	5.446	31161974	98570320	10.101	10.451
6)	B beta-BHC	4.401	4.822	16078600	57343733	11.794	10.486
7)	B delta-BHC	4.602	5.037	28058739	95525413	9.813	9.564
8)	B Heptachlo...	5.130	5.832	27457775	22119604	9.935	3.453 #
9)	A Endosulfan I	5.467	6.192	26817904	84952814	10.516	11.039
10)	B trans-Chl...	5.354	6.069	28776069	86870966	10.047	10.588
11)	B cis-Chlor...	5.412	6.140	27796536	84890290	9.981	10.561
12)	B 4,4'-DDE	5.576	6.301	48074941	149.0E6	19.308	20.901
13)	MA Dieldrin	5.708	6.450	57140624	170.2E6	20.590	21.777
14)	MA Endrin	5.965	6.670	50346928	154.0E6	21.065	23.116
15)	B Endosulfa...	6.239	6.885	44740966	147.1E6	20.259	22.166
16)	A 4,4'-DDD	6.090	6.797	41324474	134.2E6	20.983	22.147
17)	MA 4,4'-DDT	6.329	7.093	42998758	136.8E6	20.934	22.733
18)	B Endrin al...	6.407	7.009	33267287	109.6E6	20.534	21.414
19)	B Endosulfa...	6.619	7.233	38535573	131.6E6	20.613	21.564
20)	A Methoxychlor	6.876	7.557	105.0E6	341.4E6	111.528	114.881
21)	B Endrin ke...	7.110	7.706	41572942	153.5E6	21.162	21.603

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121821\
Data File : PD067424.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2021 19:39
Operator : AR\AJ
Sample : RESC007
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 01:11:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121821CLP.M
Quant Title : GC Extractables
QLast Update : Sat Dec 18 01:09:55 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

